

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017097.D
 Acq On : 01 Jul 2020 14:52
 Operator : JC/SP
 Sample : L3163-03 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-LDC-3

Quant Time: Jul 02 02:24:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	225190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	383801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	391774	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208628	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	148145	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	5.46	113	129611	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
50) Toluene-d8	8.70	98	468600	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.12	95	192402	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
Target Compounds						
16) Acetone	2.42	43	37894	36.488	ug/l	97
17) Carbon Disulfide	2.55	76	827051	132.990	ug/l	100
25) 2-Butanone	4.64	43	6111	3.531	ug/l #	88
40) Benzene	6.12	78	12573	1.180	ug/l	96
52) Toluene	8.77	92	2719	0.409	ug/l	94
68) m/p-Xylenes	10.34	106	3007	0.580	ug/l	94
95) Naphthalene	13.82	128	665177	50.276	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017097.D
Acq On : 01 Jul 2020 14:52
Operator : JC/SP
Sample : L3163-03 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-LDC-3

Quant Time: Jul 02 02:24:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

