

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\  
 Data File : VX017099.D  
 Acq On : 01 Jul 2020 15:38  
 Operator : JC/SP  
 Sample : L3129-02 20X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 1445

Quant Time: Jul 02 02:29:27 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 29 17:10:38 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.64  | 168  | 219274   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 378852   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 389041   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 206298   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 148233   | 54.84 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 109.68% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 127178   | 53.07 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 106.14% |          |
| 50) Toluene-d8              | 8.70  | 98   | 460071   | 51.92 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 103.84% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 190020   | 54.85 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 109.70% |          |
| Target Compounds            |       |      |          |       |         |          |
| 16) Acetone                 | 2.42  | 43   | 6496     | 6.424 | ug/l #  | 86       |
| 17) Carbon Disulfide        | 2.55  | 76   | 14995    | 2.499 | ug/l    | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\  
Data File : VX017099.D  
Acq On : 01 Jul 2020 15:38  
Operator : JC/SP  
Sample : L3129-02 20X  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
1445

Quant Time: Jul 02 02:29:27 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 29 17:10:38 2020  
Response via : Initial Calibration

