

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017100.D
 Acq On : 01 Jul 2020 16:00
 Operator : JC/SP
 Sample : L3129-03 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1405-1406

Quant Time: Jul 02 02:41:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	203551	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	340551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	366260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201496	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	132638	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
35) Dibromofluoromethane	5.46	113	113546	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
50) Toluene-d8	8.70	98	416218	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
62) 4-Bromofluorobenzene	11.12	95	183847	59.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.08%	
Target Compounds						
16) Acetone	2.43	43	14693	15.652	ug/l #	88
17) Carbon Disulfide	2.54	76	5915	1.075	ug/l	98
25) 2-Butanone	4.65	43	8983	5.743	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017100.D
Acq On : 01 Jul 2020 16:00
Operator : JC/SP
Sample : L3129-03 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1405-1406

Quant Time: Jul 02 02:41:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

