

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017102.D
 Acq On : 01 Jul 2020 16:46
 Operator : JC/SP
 Sample : L3163-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-LGC-3

Quant Time: Jul 02 03:10:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	228566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	389826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	412010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230888	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	151950	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
35) Dibromofluoromethane	5.46	113	128184	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
50) Toluene-d8	8.70	98	474301	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
62) 4-Bromofluorobenzene	11.12	95	214691	60.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.46%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.00	59	2858	5.317	ug/l	96
16) Acetone	2.42	43	433377	411.135	ug/l	96
17) Carbon Disulfide	2.55	76	501719	79.494	ug/l	99
25) 2-Butanone	4.64	43	21378	12.171	ug/l #	89
29) Tetrahydrofuran	5.09	42	3249	2.800	ug/l #	67
30) Chloroform	5.19	83	2880	0.638	ug/l #	81
40) Benzene	6.12	78	8916	0.824	ug/l	98
95) Naphthalene	13.82	128	51245	3.500	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017102.D
Acq On : 01 Jul 2020 16:46
Operator : JC/SP
Sample : L3163-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-LGC-3

Quant Time: Jul 02 03:10:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

