

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017107.D
 Acq On : 01 Jul 2020 18:40
 Operator : JC/SP
 Sample : L3164-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW180-SOIL-594-598

Quant Time: Jul 02 03:24:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	224466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	384257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	403505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	204409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	148219	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
35) Dibromofluoromethane	5.47	113	126204	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
50) Toluene-d8	8.70	98	470407	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
62) 4-Bromofluorobenzene	11.12	95	196984	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
Target Compounds						
16) Acetone	2.42	43	9950	9.612	ug/l	99
20) Methylene Chloride	2.84	84	25347	10.051	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017107.D
Acq On : 01 Jul 2020 18:40
Operator : JC/SP
Sample : L3164-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW180-SOIL-594-598

Quant Time: Jul 02 03:24:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

