

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070120\
 Data File : VX017118.D
 Acq On : 01 Jul 2020 23:35
 Operator : JC/SP
 Sample : L3159-05DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 07BR-20200629DL

Quant Time: Jul 02 04:09:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	226578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	370659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	383083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	144425	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	5.47	113	121603	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
50) Toluene-d8	8.70	98	449065	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	11.12	95	189769	55.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.98%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.18	85	1246	0.568	ug/l	91
4) Vinyl Chloride	1.39	62	16512	6.707	ug/l	99
21) trans-1,2-Dichloroethene	3.15	96	4061	1.632	ug/l	81
27) cis-1,2-Dichloroethene	4.57	96	292336	103.713	ug/l	91
44) Trichloroethene	7.19	130	48465	16.858	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070120\
Data File : VX017118.D
Acq On : 01 Jul 2020 23:35
Operator : JC/SP
Sample : L3159-05DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
07BR-20200629DL

Quant Time: Jul 02 04:09:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

