

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023079.D
 Acq On : 01 Jul 2021 12:55
 Operator : JC/MD
 Sample : VX0701WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBS01

Quant Time: Jul 01 14:59:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	66161	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	131107	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118777	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52556	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62110	51.73	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.46%
35) Dibromofluoromethane	5.391	113	41792	50.20	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.40%
50) Toluene-d8	8.653	98	160178	49.09	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.18%
62) 4-Bromofluorobenzene	11.085	95	60031	49.30	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.60%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	16629	17.81	ug/l	96
3) Chloromethane	1.294	50	18020	17.58	ug/l	98
4) Vinyl Chloride	1.374	62	16636	17.59	ug/l	100
5) Bromomethane	1.599	94	9201	17.66	ug/l	93
6) Chloroethane	1.685	64	10441	17.86	ug/l	92
7) Trichlorofluoromethane	1.886	101	26675	19.41	ug/l	88
8) Diethyl Ether	2.136	74	9758	18.40	ug/l	82
9) 1,1,2-Trichlorotrifluo...	2.331	101	16483	19.16	ug/l	87
10) Methyl Iodide	2.453	142	14456	16.77	ug/l	91
11) Tert butyl alcohol	2.977	59	24124	96.40	ug/l	98
12) 1,1-Dichloroethene	2.319	96	16214	18.89	ug/l	95
13) Acrolein	2.245	56	12847	82.55	ug/l	97
14) Allyl chloride	2.666	41	30775	18.87	ug/l	92
15) Acrylonitrile	3.074	53	55762	97.48	ug/l	98
16) Acetone	2.386	43	50200	94.31	ug/l	94
17) Carbon Disulfide	2.514	76	32543	16.45	ug/l	96
18) Methyl Acetate	2.709	43	26011	19.86	ug/l	90
19) Methyl tert-butyl Ether	3.123	73	60662	19.71	ug/l	96
20) Methylene Chloride	2.788	84	19855	19.04	ug/l	86
21) trans-1,2-Dichloroethene	3.099	96	16994	17.85	ug/l	95
22) Diisopropyl ether	3.769	45	66959	19.74	ug/l	# 93
23) Vinyl Acetate	3.733	43	280713	98.65	ug/l	94
24) 1,1-Dichloroethane	3.617	63	34455	19.01	ug/l	99
25) 2-Butanone	4.568	43	79981	96.62	ug/l	# 87
26) 2,2-Dichloropropane	4.483	77	24606	17.89	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	20127	19.05	ug/l	94
28) Bromochloromethane	4.903	49	16469	19.29	ug/l	# 75
29) Tetrahydrofuran	5.019	42	51969	97.61	ug/l	87
30) Chloroform	5.105	83	35179	19.30	ug/l	99
31) Cyclohexane	5.470	56	30702	18.11	ug/l	# 83
32) 1,1,1-Trichloroethane	5.397	97	29296	20.00	ug/l	# 93
36) 1,1-Dichloropropene	5.702	75	25291	18.16	ug/l	97
37) Ethyl Acetate	4.727	43	29810	18.48	ug/l	# 94
38) Carbon Tetrachloride	5.678	117	21699	18.58	ug/l	98
39) Methylcyclohexane	7.385	83	29039	17.88	ug/l	# 88
40) Benzene	6.043	78	74216	18.48	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023079.D
 Acq On : 01 Jul 2021 12:55
 Operator : JC/MD
 Sample : VX0701WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBS01

Quant Time: Jul 01 14:59:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17334	19.51	ug/l	90
42) 1,2-Dichloroethane	6.098	62	29904	20.24	ug/l	96
43) Isopropyl Acetate	6.348	43	50309	19.40	ug/l #	89
44) Trichloroethene	7.129	130	17338	18.21	ug/l	95
45) 1,2-Dichloropropane	7.433	63	20109	19.42	ug/l	96
46) Dibromomethane	7.586	93	12664	18.43	ug/l #	82
47) Bromodichloromethane	7.830	83	23112	18.52	ug/l	95
48) Methyl methacrylate	7.696	41	23462	19.55	ug/l	85
49) 1,4-Dioxane	7.665	88	9668	374.18	ug/l #	91
51) 4-Methyl-2-Pentanone	8.580	43	164229	101.75	ug/l	88
52) Toluene	8.720	92	44326	18.26	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	25712	18.09	ug/l	99
54) cis-1,3-Dichloropropene	8.372	75	28455	17.90	ug/l #	84
55) 1,1,2-Trichloroethane	9.159	97	18229	18.72	ug/l	92
56) Ethyl methacrylate	9.122	69	30805	18.83	ug/l #	73
57) 1,3-Dichloropropane	9.311	76	33351	18.84	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	80873	93.83	ug/l	93
59) 2-Hexanone	9.433	43	123074	100.51	ug/l	86
60) Dibromochloromethane	9.525	129	14878	18.11	ug/l	100
61) 1,2-Dibromoethane	9.616	107	17923	18.12	ug/l	100
64) Tetrachloroethene	9.275	164	15037	18.01	ug/l #	85
65) Chlorobenzene	10.079	112	46687	18.66	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.171	131	14741	18.24	ug/l	98
67) Ethyl Benzene	10.195	91	86614	18.80	ug/l	97
68) m/p-Xylenes	10.305	106	64191	37.53	ug/l	93
69) o-Xylene	10.646	106	30662	18.42	ug/l	92
70) Styrene	10.658	104	51589	18.67	ug/l	94
71) Bromoform	10.799	173	8189	17.43	ug/l #	98
73) Isopropylbenzene	10.963	105	83628	18.81	ug/l	99
74) N-amyl acetate	10.847	43	42066	18.13	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	29657	19.42	ug/l	98
76) 1,2,3-Trichloropropane	11.244	75	31531	22.71	ug/l	87
77) Bromobenzene	11.201	156	16831	18.20	ug/l	77
78) n-propylbenzene	11.305	91	103699	19.27	ug/l	96
79) 2-Chlorotoluene	11.366	91	62363	19.20	ug/l	94
80) 1,3,5-Trimethylbenzene	11.457	105	71283	19.52	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	6739	17.92	ug/l #	71
82) 4-Chlorotoluene	11.457	91	71246	18.98	ug/l	94
83) tert-Butylbenzene	11.719	119	63693	18.47	ug/l	87
84) 1,2,4-Trimethylbenzene	11.756	105	71471	19.45	ug/l	96
85) sec-Butylbenzene	11.896	105	88818	19.10	ug/l	98
86) p-Isopropyltoluene	12.012	119	70114	18.97	ug/l	95
87) 1,3-Dichlorobenzene	11.969	146	32863	18.42	ug/l	95
88) 1,4-Dichlorobenzene	12.042	146	33399	17.90	ug/l	93
89) n-Butylbenzene	12.335	91	67751	19.28	ug/l	94
90) Hexachloroethane	12.542	117	9635	16.76	ug/l	77
91) 1,2-Dichlorobenzene	12.341	146	32309	18.55	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	5512	18.04	ug/l	73
93) 1,2,4-Trichlorobenzene	13.591	180	17512	18.13	ug/l	96
94) Hexachlorobutadiene	13.731	225	7027	18.57	ug/l	100
95) Naphthalene	13.780	128	71905	18.06	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	18472	18.99	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023079.D
Acq On : 01 Jul 2021 12:55
Operator : JC/MD
Sample : VX0701WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Quant Time: Jul 01 14:59:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023079.D
 Acq On : 01 Jul 2021 12:55
 Operator : JC/MD
 Sample : VX0701WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBS01

Quant Time: Jul 01 14:59:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion:168 Resp: 66161
Ion Ratio Lower Upper
168 100
99 73.4 45.8 68.6#




#2
Dichlorodifluoromethane
Concen: 17.81 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55


Tgt Ion: 85 Resp: 16629
Ion Ratio Lower Upper
85 100
87 35.2 16.4 49.2




Abundance Scan 34 (1.294 min): VX022839.D\data.ms (-30)



Abundance Scan 34 (1.294 min): VX023079.D\data.ms




Abundance Scan 34 (1.294 min): VX023079.D\data.ms (-1)




#3
Chloromethane
Concen: 17.58 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55


Tgt Ion: 50 Resp: 18020
Ion Ratio Lower Upper
50 100
52 32.0 26.4 39.6




Abundance Scan 47 (1.374 min): VX022839.D\data.ms (-43)



Abundance Scan 47 (1.374 min): VX023079.D\data.ms




Abundance Scan 47 (1.374 min): VX023079.D\data.ms (-1)




#4
Vinyl Chloride
Concen: 17.59 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 62 Resp: 16636
Ion Ratio Lower Upper
62 100
64 31.9 25.7 38.5







Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Abundance Scan 84 (1.599 min): VX022839.D\data.ms (-78)






#5
Bromomethane
Concen: 17.66 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55


Tgt Ion: 94 Resp: 9201
Ion Ratio Lower Upper
94 100
96 88.9 76.4 114.6



#6
Chloroethane
Concen: 17.86 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.007 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 64 Resp: 10441
Ion Ratio Lower Upper
64 100
66 36.1 25.4 38.0









#7
Trichlorofluoromethane
Concen: 19.41 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:101 Resp: 26675
Ion Ratio Lower Upper
101 100
103 56.7 53.2 79.8



#8
Diethyl Ether
Concen: 18.40 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 74 Resp: 9758
Ion Ratio Lower Upper
74 100
45 105.7 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.16 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:	101	Resp:	16483
Ion	Ratio	Lower	Upper
101	100		
85	50.1	34.9	52.3
151	63.9	61.4	92.2



#10
Methyl Iodide
Concen: 16.77 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:	142	Resp:	14456
Ion	Ratio	Lower	Upper
142	100		
127	49.0	33.3	49.9
141	14.6	11.4	17.0









#11
Tert butyl alcohol
Concen: 96.40 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 59 Resp: 24124
Ion Ratio Lower Upper
59 100
57 9.2 7.8 11.8



#12
1,1-Dichloroethene
Concen: 18.89 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.001 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 96 Resp: 16214
Ion Ratio Lower Upper
96 100
61 158.4 124.6 187.0
98 52.8 50.6 76.0









#13
Acrolein
Concen: 82.55 ug/l
RT: 2.245 min Scan# 190
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 56 Resp: 12847
Ion Ratio Lower Upper
56 100
55 73.0 56.4 84.6



#14
Allyl chloride
Concen: 18.87 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 41 Resp: 30775
Ion Ratio Lower Upper
41 100
39 68.2 51.6 77.4
76 30.5 30.5 45.7









#15
Acrylonitrile
Concen: 97.48 ug/l
RT: 3.074 min Scan# 326
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:	53	Resp:	55762
Ion	Ratio	Lower	Upper
53	100		
52	83.6	65.3	97.9
51	36.8	28.2	42.2



#16
Acetone
Concen: 94.31 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:	43	Resp:	50200
Ion	Ratio	Lower	Upper
43	100		
58	31.1	27.8	41.8









#17
Carbon Disulfide
Concen: 16.45 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 76 Resp: 32543
Ion Ratio Lower Upper
76 100
78 10.3 7.2 10.8



#18
Methyl Acetate
Concen: 19.86 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 43 Resp: 26011
Ion Ratio Lower Upper
43 100
74 22.0 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 19.71 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 73 Resp: 60662
Ion Ratio Lower Upper
73 100
57 23.1 17.1 25.7



#20
Methylene Chloride
Concen: 19.04 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 84 Resp: 19855
Ion Ratio Lower Upper
84 100
49 129.6 88.6 132.8
51 38.1 27.0 40.6
86 55.4 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 17.85 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 96 Resp: 16994
Ion Ratio Lower Upper
96 100
61 142.6 107.7 161.5
98 61.4 50.4 75.6



#22
Diisopropyl ether
Concen: 19.74 ug/l
RT: 3.769 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 45 Resp: 66959
Ion Ratio Lower Upper
45 100
43 76.8 58.0 87.0
87 24.6 25.4 38.0#
59 12.4 10.2 15.2









#23
 Vinyl Acetate
 Concen: 98.65 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. -0.000 min
 Lab File: VX023079.D
 Acq: 01 Jul 2021 12:55




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBS01


Tgt Ion: 43 Resp: 280713
 Ion Ratio Lower Upper
 43 100
 86 10.1 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 19.01 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX023079.D
 Acq: 01 Jul 2021 12:55

Tgt Ion: 63 Resp: 34455
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.8 11.2
 100 4.8 2.5 7.5









#25
2-Butanone
Concen: 96.62 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 43 Resp: 79981
Ion Ratio Lower Upper
43 100
72 22.6 23.7 35.5#



#26
2,2-Dichloropropane
Concen: 17.89 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 77 Resp: 24606
Ion Ratio Lower Upper
77 100
97 22.0 11.7 35.1









#27
cis-1,2-Dichloroethene
Concen: 19.05 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion: 96 Resp: 20127




Ion	Ratio	Lower	Upper
96	100		
61	153.2	0.0	285.6
98	63.5	0.0	129.6




#28
Bromochloromethane
Concen: 19.29 ug/l
RT: 4.903 min Scan# 626
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 49 Resp: 16469

Ion	Ratio	Lower	Upper
49	100		
129	0.6	0.0	4.0
130	62.4	68.2	102.2#









#29
Tetrahydrofuran
Concen: 97.61 ug/l
RT: 5.019 min Scan# 645
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:	42	Resp:	51969
Ion	Ratio	Lower	Upper
42	100		
72	43.0	41.7	62.5
71	38.5	38.4	57.6



#30
Chloroform
Concen: 19.30 ug/l
RT: 5.105 min Scan# 659
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:	83	Resp:	35179
Ion	Ratio	Lower	Upper
83	100		
85	64.1	52.0	78.0









#31
Cyclohexane
Concen: 18.11 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:	56	Resp:	30702
Ion	Ratio	Lower	Upper
56	100		
69	31.1	25.3	37.9
84	67.3	71.4	107.0#



#32
1,1,1-Trichloroethane
Concen: 20.00 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:	97	Resp:	29296
Ion	Ratio	Lower	Upper
97	100		
99	63.6	51.5	77.3
61	52.6	34.2	51.4#





#33

1,2-Dichloroethane-d4

Concen: 51.73 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX023079.D





Acq: 01 Jul 2021 12:55

Tgt Ion: 65 Resp: 62110

Ion Ratio Lower Upper

65 100

67 52.5 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55




Tgt Ion: 114 Resp: 131107


Ion Ratio Lower Upper

114 100

63 23.2 0.0 40.8

88 16.7 0.0 32.8





#35

Dibromofluoromethane

Concen: 50.20 ug/l

RT: 5.391 min Scan# 706





Delta R.T. -0.006 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55

Tgt Ion:	113	Resp:	41792
Ion	Ratio	Lower	Upper
113	100		
111	103.6	81.9	122.9
192	14.3	16.7	25.1#

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01



#36

1,1-Dichloropropene

Concen: 18.16 ug/l




RT: 5.702 min Scan# 757


Delta R.T. 0.000 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55

Tgt Ion:	75	Resp:	25291
Ion	Ratio	Lower	Upper
75	100		
110	33.7	18.1	54.4
77	31.4	25.3	37.9





#37

Ethyl Acetate

Concen: 18.48 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55





Instrument :

MSVOA_X

ClientSampleId :

VX0701WBS01

Tgt Ion:	43	Resp:	29810
Ion	Ratio	Lower	Upper
43	100		
61	14.2	12.8	19.2
70	10.1	10.5	15.7#



#38

Carbon Tetrachloride

Concen: 18.58 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.012 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55

Tgt Ion:	117	Resp:	21699
Ion	Ratio	Lower	Upper
117	100		
119	93.8	76.2	114.2
121	28.5	24.8	37.2





#39

Methylcyclohexane

Concen: 17.88 ug/l

RT: 7.385 min Scan# 1033





Delta R.T. -0.000 min

Lab File: VX023079.D

Acq: 01 Jul 2021 12:55

Tgt Ion: 83 Resp: 29039

Ion	Ratio	Lower	Upper
83	100		
55	90.0	58.9	88.3#
98	48.5	38.1	57.1



#40

Benzene

Concen: 18.48 ug/l

RT: 6.043 min Scan# 813




Delta R.T. -0.007 min


Lab File: VX023079.D

Acq: 01 Jul 2021 12:55

Tgt Ion: 78 Resp: 74216

Ion	Ratio	Lower	Upper
78	100		
77	23.2	18.9	28.3









#41
Methacrylonitrile
Concen: 19.51 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion: 41 Resp: 17334




Ion	Ratio	Lower	Upper
41	100		
39	58.0	42.5	63.7
67	67.8	65.0	97.6
52	29.7	25.3	37.9




#42
1,2-Dichloroethane
Concen: 20.24 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 62 Resp: 29904

Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	20.2









#43
Isopropyl Acetate
Concen: 19.40 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 43 Resp: 50309
Ion Ratio Lower Upper
43 100
61 17.8 18.2 27.4#
87 11.0 12.5 18.7#



#44
Trichloroethene
Concen: 18.21 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:130 Resp: 17338
Ion Ratio Lower Upper
130 100
95 103.9 0.0 198.0









#45
1,2-Dichloropropane
Concen: 19.42 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 63 Resp: 20109
Ion Ratio Lower Upper
63 100
65 30.0 25.7 38.5



#46
Dibromomethane
Concen: 18.43 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 93 Resp: 12664
Ion Ratio Lower Upper
93 100
95 86.5 66.8 100.2
174 73.6 83.1 124.7#









#47
Bromodichloromethane
Concen: 18.52 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion: 83 Resp: 23112




Ion	Ratio	Lower	Upper
83	100		
85	59.9	51.4	77.0
127	7.2	6.8	10.2




#48
Methyl methacrylate
Concen: 19.55 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 41 Resp: 23462

Ion	Ratio	Lower	Upper
41	100		
69	78.0	76.0	114.0
39	63.7	44.6	67.0









#49
1,4-Dioxane
Concen: 374.18 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampled :
VX0701WBS01


Tgt Ion: 88 Resp: 9668
Ion Ratio Lower Upper
88 100
43 37.5 23.2 34.8#
58 74.1 55.0 82.4



#50
Toluene-d8
Concen: 49.09 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 98 Resp: 160178
Ion Ratio Lower Upper
98 100
100 65.8 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 101.75 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 43 Resp: 164229
Ion Ratio Lower Upper
43 100
58 36.1 35.4 53.0



#52
Toluene
Concen: 18.26 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 92 Resp: 44326
Ion Ratio Lower Upper
92 100
91 172.2 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 18.09 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 75 Resp: 25712
Ion Ratio Lower Upper
75 100
77 31.9 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 17.90 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 75 Resp: 28455
Ion Ratio Lower Upper
75 100
77 33.2 25.7 38.5
39 61.3 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 18.72 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion: 97 Resp: 18229




Ion	Ratio	Lower	Upper
97	100		
83	95.9	69.3	103.9
85	58.9	44.3	66.5
99	57.9	50.3	75.5




#56
Ethyl methacrylate
Concen: 18.83 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 69 Resp: 30805

Ion	Ratio	Lower	Upper
69	100		
41	75.0	46.8	70.2#
39	50.7	24.7	37.1#









#57
1,3-Dichloropropane
Concen: 18.84 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 76 Resp: 33351
Ion Ratio Lower Upper
76 100
78 31.9 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 93.83 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 63 Resp: 80873
Ion Ratio Lower Upper
63 100
106 25.7 23.5 35.3









#59
2-Hexanone
Concen: 100.51 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 43 Resp: 123074
Ion Ratio Lower Upper
43 100
58 50.9 30.9 92.8



#60
Dibromochloromethane
Concen: 18.11 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.001 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 129 Resp: 14878
Ion Ratio Lower Upper
129 100
127 76.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 18.12 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:107 Resp: 17923
Ion Ratio Lower Upper
107 100
109 94.3 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 49.30 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 95 Resp: 60031
Ion Ratio Lower Upper
95 100
174 62.3 0.0 154.6
176 58.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:117 Resp: 118777
Ion Ratio Lower Upper
117 100
82 63.4 47.4 71.2
119 33.4 26.6 39.8



#64
Tetrachloroethene
Concen: 18.01 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:164 Resp: 15037
Ion Ratio Lower Upper
164 100
166 123.5 103.4 155.2
129 116.6 70.8 106.2#
131 103.4 72.5 108.7









#65
Chlorobenzene
Concen: 18.66 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:112 Resp: 46687
Ion Ratio Lower Upper
112 100
114 32.0 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 18.24 ug/l
RT: 10.171 min Scan# 1490
Delta R.T. 0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:131 Resp: 14741
Ion Ratio Lower Upper
131 100
133 95.3 47.7 143.1
119 67.9 32.1 96.5









#67
Ethyl Benzene
Concen: 18.80 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 91 Resp: 86614
Ion Ratio Lower Upper
91 100
106 29.3 24.6 37.0



#68
m/p-Xylenes
Concen: 37.53 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 106 Resp: 64191
Ion Ratio Lower Upper
106 100
91 214.4 162.8 244.2









#69
o-Xylene
Concen: 18.42 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:106 Resp: 30662
Ion Ratio Lower Upper
106 100
91 228.9 108.1 324.2



#70
Styrene
Concen: 18.67 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:104 Resp: 51589
Ion Ratio Lower Upper
104 100
78 58.3 40.0 60.0
103 54.0 43.4 65.0









#71
Bromoform
Concen: 17.43 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:173 Resp: 8189
Ion Ratio Lower Upper
173 100
175 46.5 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:152 Resp: 52556
Ion Ratio Lower Upper
152 100
115 78.8 41.1 123.3
150 162.5 0.0 349.0









#73
Isopropylbenzene
Concen: 18.81 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:105 Resp: 83628
Ion Ratio Lower Upper
105 100
120 25.9 13.2 39.6



#74
N-ethyl acetate
Concen: 18.13 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 43 Resp: 42066
Ion Ratio Lower Upper
43 100
70 40.6 39.0 58.6
55 23.6 21.5 32.3
61 22.2 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 19.42 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 83 Resp: 29657
Ion Ratio Lower Upper
83 100
131 9.3 5.1 15.2
85 61.9 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 22.71 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 75 Resp: 31531
Ion Ratio Lower Upper
75 100
77 30.6 19.1 57.5









#77
Bromobenzene
Concen: 18.20 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:156 Resp: 16831
Ion Ratio Lower Upper
156 100
77 214.4 82.8 248.6
158 96.3 49.1 147.4



#78
n-propylbenzene
Concen: 19.27 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 91 Resp: 103699
Ion Ratio Lower Upper
91 100
120 21.0 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.20 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 91 Resp: 62363
Ion Ratio Lower Upper
91 100
126 30.0 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 19.52 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 105 Resp: 71283
Ion Ratio Lower Upper
105 100
120 45.8 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 17.92 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion: 75 Resp: 6739
Ion Ratio Lower Upper
75 100
53 133.0 74.5 111.7#
89 44.6 37.9 56.9



#82
4-Chlorotoluene
Concen: 18.98 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 91 Resp: 71246
Ion Ratio Lower Upper
91 100
126 26.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 18.47 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion:119 Resp: 63693




Ion	Ratio	Lower	Upper
119	100		
91	68.6	27.9	83.7
134	23.5	11.3	33.9




#84
1,2,4-Trimethylbenzene
Concen: 19.45 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:105 Resp: 71471

Ion	Ratio	Lower	Upper
105	100		
120	42.4	22.4	67.3









#85
sec-Butylbenzene
Concen: 19.10 ug/l
RT: 11.896 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:105 Resp: 88818
Ion Ratio Lower Upper
105 100
134 18.9 10.0 29.8



#86
p-Isopropyltoluene
Concen: 18.97 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:119 Resp: 70114
Ion Ratio Lower Upper
119 100
134 25.8 12.9 38.7
91 28.0 11.7 35.0









#87

1,3-Dichlorobenzene
Concen: 18.42 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01




Tgt Ion:146 Resp: 32863
Ion Ratio Lower Upper
146 100
111 47.1 20.2 60.5
148 64.4 31.9 95.7






#88

1,4-Dichlorobenzene
Concen: 17.90 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55



Tgt Ion:146 Resp: 33399
Ion Ratio Lower Upper
146 100
111 48.0 19.6 58.8
148 65.6 31.6 95.0









#89
n-Butylbenzene
Concen: 19.28 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion: 91 Resp: 67751




Ion	Ratio	Lower	Upper
91	100		
92	51.4	27.4	82.0
134	22.1	13.2	39.5




#90
Hexachloroethane
Concen: 16.76 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 117 Resp: 9635

Ion	Ratio	Lower	Upper
117	100		
201	60.7	40.4	121.2









#91
1,2-Dichlorobenzene
Concen: 18.55 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:146 Resp: 32309
Ion Ratio Lower Upper
146 100
111 48.2 20.8 62.2
148 65.1 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.04 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion: 75 Resp: 5512
Ion Ratio Lower Upper
75 100
155 65.6 45.3 135.9
157 84.9 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 18.13 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01


Tgt Ion:180 Resp: 17512
Ion Ratio Lower Upper
180 100
182 94.6 47.7 143.1
145 36.6 15.0 45.0



#94
Hexachlorobutadiene
Concen: 18.57 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:225 Resp: 7027
Ion Ratio Lower Upper
225 100
223 62.2 31.1 93.2
227 62.8 31.6 95.0









#95
Naphthalene
Concen: 18.06 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBS01

Tgt Ion:	128	Resp:	71905
Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.3	15.5
129	11.1	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.99 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX023079.D
Acq: 01 Jul 2021 12:55

Tgt Ion:	180	Resp:	18472
Ion	Ratio	Lower	Upper
180	100		
182	92.2	47.6	142.9
145	37.5	16.1	48.3

