

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023084.D
 Acq On : 01 Jul 2021 14:52
 Operator : JC/MD
 Sample : VX0701WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBSD01

Quant Time: Jul 01 16:05:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65492	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	133636	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	121228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51307	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	67180	56.52	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.04%
35) Dibromofluoromethane	5.397	113	41315	48.68	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.36%
50) Toluene-d8	8.653	98	160191	48.16	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.32%
62) 4-Bromofluorobenzene	11.085	95	59866	48.23	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.46%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	16378	17.72	ug/l	98
3) Chloromethane	1.294	50	18807	18.54	ug/l	98
4) Vinyl Chloride	1.373	62	17276	18.45	ug/l	99
5) Bromomethane	1.599	94	8225	15.95	ug/l	94
6) Chloroethane	1.678	64	10925	18.88	ug/l	96
7) Trichlorofluoromethane	1.879	101	25602	18.82	ug/l	95
8) Diethyl Ether	2.135	74	9639	18.36	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.324	101	15450	18.14	ug/l	88
10) Methyl Iodide	2.452	142	15063	17.45	ug/l	92
11) Tert butyl alcohol	2.983	59	24557	99.14	ug/l	99
12) 1,1-Dichloroethene	2.318	96	15518	18.26	ug/l	85
13) Acrolein	2.239	56	13148	85.34	ug/l	92
14) Allyl chloride	2.666	41	30887	19.13	ug/l	# 91
15) Acrylonitrile	3.074	53	58452	103.22	ug/l	98
16) Acetone	2.392	43	51326	97.41	ug/l	94
17) Carbon Disulfide	2.507	76	31214	16.05	ug/l	99
18) Methyl Acetate	2.715	43	27320	21.07	ug/l	91
19) Methyl tert-butyl Ether	3.123	73	61321	20.13	ug/l	99
20) Methylene Chloride	2.794	84	19163	18.57	ug/l	# 86
21) trans-1,2-Dichloroethene	3.093	96	16837	17.86	ug/l	91
22) Diisopropyl ether	3.769	45	67817	20.20	ug/l	# 87
23) Vinyl Acetate	3.733	43	288274	102.35	ug/l	# 93
24) 1,1-Dichloroethane	3.617	63	36070	20.11	ug/l	97
25) 2-Butanone	4.568	43	82644	100.86	ug/l	89
26) 2,2-Dichloropropane	4.483	77	25135	18.46	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	20346	19.45	ug/l	90
28) Bromochloromethane	4.903	49	17783	21.04	ug/l	# 71
29) Tetrahydrofuran	5.019	42	53157	100.86	ug/l	88
30) Chloroform	5.104	83	34607	19.18	ug/l	97
31) Cyclohexane	5.470	56	29654	17.67	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	27308	18.84	ug/l	93
36) 1,1-Dichloropropene	5.702	75	25142	17.71	ug/l	94
37) Ethyl Acetate	4.726	43	32777	19.93	ug/l	# 93
38) Carbon Tetrachloride	5.684	117	20954	17.60	ug/l	96
39) Methylcyclohexane	7.385	83	28167	17.01	ug/l	92
40) Benzene	6.043	78	74684	18.25	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023084.D
 Acq On : 01 Jul 2021 14:52
 Operator : JC/MD
 Sample : VX0701WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBSD01

Quant Time: Jul 01 16:05:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17596	19.43	ug/l	90
42) 1,2-Dichloroethane	6.098	62	31274	20.77	ug/l	94
43) Isopropyl Acetate	6.348	43	51776	19.59	ug/l #	92
44) Trichloroethene	7.135	130	17789	18.33	ug/l	95
45) 1,2-Dichloropropane	7.439	63	20175	19.11	ug/l	98
46) Dibromomethane	7.586	93	12893	18.41	ug/l #	83
47) Bromodichloromethane	7.830	83	23162	18.25	ug/l	91
48) Methyl methacrylate	7.702	41	24033	19.64	ug/l	85
49) 1,4-Dioxane	7.671	88	8730	331.49	ug/l #	88
51) 4-Methyl-2-Pentanone	8.579	43	168368	102.34	ug/l	90
52) Toluene	8.720	92	46338	18.73	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	24732	17.27	ug/l	95
54) cis-1,3-Dichloropropene	8.372	75	27998	17.37	ug/l #	87
55) 1,1,2-Trichloroethane	9.159	97	18970	19.11	ug/l	98
56) Ethyl methacrylate	9.122	69	31346	18.80	ug/l #	80
57) 1,3-Dichloropropane	9.311	76	33507	18.57	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.250	63	84730	96.30	ug/l	93
59) 2-Hexanone	9.433	43	127990	102.54	ug/l	86
60) Dibromochloromethane	9.524	129	14007	17.00	ug/l	99
61) 1,2-Dibromoethane	9.616	107	18915	18.76	ug/l	99
64) Tetrachloroethene	9.274	164	16246	19.07	ug/l	92
65) Chlorobenzene	10.085	112	45623	17.87	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	15265	18.50	ug/l	97
67) Ethyl Benzene	10.195	91	87411	18.59	ug/l	100
68) m/p-Xylenes	10.305	106	63109	36.15	ug/l	92
69) o-Xylene	10.646	106	31407	18.48	ug/l	93
70) Styrene	10.658	104	51674	18.32	ug/l	94
71) Bromoform	10.805	173	7913	16.74	ug/l #	95
73) Isopropylbenzene	10.963	105	82657	19.04	ug/l	98
74) N-amyl acetate	10.847	43	45516	19.87	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	11.219	83	28448	19.08	ug/l	97
76) 1,2,3-Trichloropropane	11.244	75	33004	24.35	ug/l	91
77) Bromobenzene	11.201	156	15833	17.54	ug/l	69
78) n-propylbenzene	11.305	91	100456	19.12	ug/l	95
79) 2-Chlorotoluene	11.366	91	59840	18.87	ug/l	94
80) 1,3,5-Trimethylbenzene	11.457	105	69465	19.49	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.024	75	7119	18.85	ug/l #	71
82) 4-Chlorotoluene	11.457	91	68913	18.80	ug/l	94
83) tert-Butylbenzene	11.719	119	62359	18.52	ug/l	86
84) 1,2,4-Trimethylbenzene	11.756	105	69797	19.45	ug/l	95
85) sec-Butylbenzene	11.896	105	86024	18.95	ug/l	97
86) p-Isopropyltoluene	12.012	119	67747	18.78	ug/l	94
87) 1,3-Dichlorobenzene	11.975	146	31465	18.07	ug/l	93
88) 1,4-Dichlorobenzene	12.042	146	32548	17.87	ug/l	94
89) n-Butylbenzene	12.335	91	64895	18.92	ug/l	97
90) Hexachloroethane	12.542	117	9294	16.61	ug/l	66
91) 1,2-Dichlorobenzene	12.341	146	32385	19.05	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.944	75	5427	18.16	ug/l	72
93) 1,2,4-Trichlorobenzene	13.591	180	16970	18.00	ug/l	96
94) Hexachlorobutadiene	13.725	225	6686	18.10	ug/l	94
95) Naphthalene	13.780	128	72281	18.54	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	17714	18.65	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023084.D
Acq On : 01 Jul 2021 14:52
Operator : JC/MD
Sample : VX0701WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Quant Time: Jul 01 16:05:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023084.D
 Acq On : 01 Jul 2021 14:52
 Operator : JC/MD
 Sample : VX0701WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBSD01

Quant Time: Jul 01 16:05:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion:168 Resp: 65492
Ion Ratio Lower Upper
168 100
99 71.4 45.8 68.6#




#2
Dichlorodifluoromethane
Concen: 17.72 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 85 Resp: 16378
Ion Ratio Lower Upper
85 100
87 33.7 16.4 49.2







Abundance Scan 34 (1.294 min): VX022839.D\data.ms (-30)






#3
Chloromethane
Concen: 18.54 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52


Tgt Ion: 50 Resp: 18807
Ion Ratio Lower Upper
50 100
52 32.0 26.4 39.6



#4
Vinyl Chloride
Concen: 18.45 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 62 Resp: 17276
Ion Ratio Lower Upper
62 100
64 32.9 25.7 38.5









#5
Bromomethane
Concen: 15.95 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 94 Resp: 8225
Ion Ratio Lower Upper
94 100
96 101.7 76.4 114.6



#6
Chloroethane
Concen: 18.88 ug/l
RT: 1.678 min Scan# 97
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 64 Resp: 10925
Ion Ratio Lower Upper
64 100
66 29.6 25.4 38.0









#7
Trichlorofluoromethane
Concen: 18.82 ug/l
RT: 1.879 min Scan# 130
Delta R.T. -0.007 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:101 Resp: 25602
Ion Ratio Lower Upper
101 100
103 70.1 53.2 79.8



#8
Diethyl Ether
Concen: 18.36 ug/l
RT: 2.135 min Scan# 172
Delta R.T. -0.007 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 74 Resp: 9639
Ion Ratio Lower Upper
74 100
45 111.7 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.14 ug/l
RT: 2.324 min Scan# 203
Delta R.T. -0.007 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:101 Resp: 15450
Ion Ratio Lower Upper
101 100
85 49.6 34.9 52.3
151 65.5 61.4 92.2



#10
Methyl Iodide
Concen: 17.45 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:142 Resp: 15063
Ion Ratio Lower Upper
142 100
127 48.1 33.3 49.9
141 14.4 11.4 17.0









#11

Tert butyl alcohol
Concen: 99.14 ug/l
RT: 2.983 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01




Tgt Ion: 59 Resp: 24557
Ion Ratio Lower Upper
59 100
57 9.6 7.8 11.8




#12

1,1-Dichloroethene
Concen: 18.26 ug/l
RT: 2.318 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 96 Resp: 15518
Ion Ratio Lower Upper
96 100
61 180.4 124.6 187.0
98 67.0 50.6 76.0









#13
Acrolein
Concen: 85.34 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 56 Resp: 13148
Ion Ratio Lower Upper
56 100
55 77.0 56.4 84.6



#14
Allyl chloride
Concen: 19.13 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 41 Resp: 30887
Ion Ratio Lower Upper
41 100
39 67.6 51.6 77.4
76 27.2 30.5 45.7#









#15
Acrylonitrile
Concen: 103.22 ug/l
RT: 3.074 min Scan# 326
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion: 53 Resp: 58452




Ion	Ratio	Lower	Upper
53	100		
52	83.1	65.3	97.9
51	36.5	28.2	42.2




#16
Acetone
Concen: 97.41 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 43 Resp: 51326

Ion	Ratio	Lower	Upper
43	100		
58	31.4	27.8	41.8









#17
Carbon Disulfide
Concen: 16.05 ug/l
RT: 2.507 min Scan# 233
Delta R.T. -0.007 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 76 Resp: 31214
Ion Ratio Lower Upper
76 100
78 9.2 7.2 10.8



#18
Methyl Acetate
Concen: 21.07 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 43 Resp: 27320
Ion Ratio Lower Upper
43 100
74 22.7 21.9 32.9









#19

Methyl tert-butyl Ether
Concen: 20.13 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01




Tgt Ion: 73 Resp: 61321
Ion Ratio Lower Upper
73 100
57 22.1 17.1 25.7




#20

Methylene Chloride
Concen: 18.57 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 84 Resp: 19163
Ion Ratio Lower Upper
84 100
49 132.9 88.6 132.8#
51 40.1 27.0 40.6
86 65.8 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 17.86 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 96 Resp: 16837
Ion Ratio Lower Upper
96 100
61 147.8 107.7 161.5
98 66.8 50.4 75.6



#22
Diisopropyl ether
Concen: 20.20 ug/l
RT: 3.769 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 45 Resp: 67817
Ion Ratio Lower Upper
45 100
43 84.7 58.0 87.0
87 24.0 25.4 38.0#
59 11.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 102.35 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. -0.000 min
 Lab File: VX023084.D
 Acq: 01 Jul 2021 14:52




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBSD01


Tgt Ion: 43 Resp: 288274
 Ion Ratio Lower Upper
 43 100
 86 9.7 9.9 14.9#



#24
 1,1-Dichloroethane
 Concen: 20.11 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX023084.D
 Acq: 01 Jul 2021 14:52

Tgt Ion: 63 Resp: 36070
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.8 11.2
 100 3.9 2.5 7.5









#25
2-Butanone
Concen: 100.86 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 43 Resp: 82644
Ion Ratio Lower Upper
43 100
72 23.8 23.7 35.5



#26
2,2-Dichloropropane
Concen: 18.46 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 77 Resp: 25135
Ion Ratio Lower Upper
77 100
97 22.2 11.7 35.1









#27
cis-1,2-Dichloroethene
Concen: 19.45 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion: 96 Resp: 20346




Ion	Ratio	Lower	Upper
96	100		
61	158.4	0.0	285.6
98	61.4	0.0	129.6




#28
Bromochloromethane
Concen: 21.04 ug/l
RT: 4.903 min Scan# 626
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 49 Resp: 17783

Ion	Ratio	Lower	Upper
49	100		
129	0.8	0.0	4.0
130	58.5	68.2	102.2#









#29
Tetrahydrofuran
Concen: 100.86 ug/l
RT: 5.019 min Scan# 645
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:	42	Resp:	53157
Ion	Ratio	Lower	Upper
42	100		
72	42.9	41.7	62.5
71	40.6	38.4	57.6



#30
Chloroform
Concen: 19.18 ug/l
RT: 5.104 min Scan# 659
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:	83	Resp:	34607
Ion	Ratio	Lower	Upper
83	100		
85	62.4	52.0	78.0









#31
Cyclohexane
Concen: 17.67 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 56 Resp: 29654
Ion Ratio Lower Upper
56 100
69 27.0 25.3 37.9
84 83.5 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 18.84 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 97 Resp: 27308
Ion Ratio Lower Upper
97 100
99 67.5 51.5 77.3
61 50.8 34.2 51.4









#33

1,2-Dichloroethane-d4
Concen: 56.52 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01




Tgt Ion: 65 Resp: 67180
Ion Ratio Lower Upper
65 100
67 50.1 0.0 105.4




#34

1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 114 Resp: 133636
Ion Ratio Lower Upper
114 100
63 24.8 0.0 40.8
88 19.3 0.0 32.8









#35
Dibromofluoromethane
Concen: 48.68 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:113 Resp: 41315
Ion Ratio Lower Upper
113 100
111 103.8 81.9 122.9
192 14.6 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 17.71 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 75 Resp: 25142
Ion Ratio Lower Upper
75 100
110 31.3 18.1 54.4
77 30.2 25.3 37.9





#37

Ethyl Acetate

Concen: 19.93 ug/l

RT: 4.726 min Scan# 597

Delta R.T. -0.001 min

Lab File: VX023084.D

Acq: 01 Jul 2021 14:52

Tgt Ion: 43 Resp: 32777

Ion Ratio Lower Upper

43 100

61 13.5 12.8 19.2





70 9.5 10.5 15.7#

Instrument :

MSVOA_X

ClientSampleId :

VX0701WBSD01



#38

Carbon Tetrachloride

Concen: 17.60 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX023084.D

Acq: 01 Jul 2021 14:52




Tgt Ion: 117 Resp: 20954


Ion Ratio Lower Upper

117 100

119 91.0 76.2 114.2

121 29.9 24.8 37.2





#39

Methylcyclohexane

Concen: 17.01 ug/l

RT: 7.385 min Scan# 1033





Delta R.T. -0.000 min

Lab File: VX023084.D

Acq: 01 Jul 2021 14:52

Tgt Ion: 83 Resp: 28167

Ion	Ratio	Lower	Upper
83	100		
55	82.8	58.9	88.3
98	49.8	38.1	57.1



#40

Benzene

Concen: 18.25 ug/l

RT: 6.043 min Scan# 813




Delta R.T. -0.007 min


Lab File: VX023084.D

Acq: 01 Jul 2021 14:52

Tgt Ion: 78 Resp: 74684

Ion	Ratio	Lower	Upper
78	100		
77	22.4	18.9	28.3









#41
Methacrylonitrile
Concen: 19.43 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion: 41 Resp: 17596




Ion	Ratio	Lower	Upper
41	100		
39	58.4	42.5	63.7
67	69.3	65.0	97.6
52	29.7	25.3	37.9




#42
1,2-Dichloroethane
Concen: 20.77 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 62 Resp: 31274

Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	20.2









#43
Isopropyl Acetate
Concen: 19.59 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 43 Resp: 51776
Ion Ratio Lower Upper
43 100
61 18.6 18.2 27.4
87 12.3 12.5 18.7#



#44
Trichloroethene
Concen: 18.33 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:130 Resp: 17789
Ion Ratio Lower Upper
130 100
95 104.3 0.0 198.0









#45
1,2-Dichloropropane
Concen: 19.11 ug/l
RT: 7.439 min Scan# 1042
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 63 Resp: 20175
Ion Ratio Lower Upper
63 100
65 30.8 25.7 38.5



#46
Dibromomethane
Concen: 18.41 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 93 Resp: 12893
Ion Ratio Lower Upper
93 100
95 83.7 66.8 100.2
174 73.7 83.1 124.7#









#47
Bromodichloromethane
Concen: 18.25 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:	83	Resp:	23162
Ion	Ratio	Lower	Upper
83	100		
85	57.1	51.4	77.0
127	6.8	6.8	10.2



#48
Methyl methacrylate
Concen: 19.64 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:	41	Resp:	24033
Ion	Ratio	Lower	Upper
41	100		
69	78.1	76.0	114.0
39	62.9	44.6	67.0









#49

1,4-Dioxane
Concen: 331.49 ug/l
RT: 7.671 min Scan# 1080
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01




Tgt Ion: 88 Resp: 8730
Ion Ratio Lower Upper
88 100
43 39.3 23.2 34.8#
58 76.0 55.0 82.4




#50

Toluene-d8
Concen: 48.16 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 98 Resp: 160191
Ion Ratio Lower Upper
98 100
100 65.4 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 102.34 ug/l
RT: 8.579 min Scan# 1229
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 43 Resp: 168368
Ion Ratio Lower Upper
43 100
58 37.5 35.4 53.0



#52
Toluene
Concen: 18.73 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 92 Resp: 46338
Ion Ratio Lower Upper
92 100
91 169.9 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 17.27 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 75 Resp: 24732
Ion Ratio Lower Upper
75 100
77 28.5 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 17.37 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 75 Resp: 27998
Ion Ratio Lower Upper
75 100
77 31.9 25.7 38.5
39 58.7 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 19.11 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion: 97 Resp: 18970




Ion	Ratio	Lower	Upper
97	100		
83	83.1	69.3	103.9
85	55.3	44.3	66.5
99	64.7	50.3	75.5




#56
Ethyl methacrylate
Concen: 18.80 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 69 Resp: 31346

Ion	Ratio	Lower	Upper
69	100		
41	70.6	46.8	70.2#
39	46.2	24.7	37.1#









#57
1,3-Dichloropropane
Concen: 18.57 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 76 Resp: 33507
Ion Ratio Lower Upper
76 100
78 30.9 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 96.30 ug/l
RT: 8.250 min Scan# 1175
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 63 Resp: 84730
Ion Ratio Lower Upper
63 100
106 25.6 23.5 35.3









#59
2-Hexanone
Concen: 102.54 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 43 Resp: 127990
Ion Ratio Lower Upper
43 100
58 51.2 30.9 92.8



#60
Dibromochloromethane
Concen: 17.00 ug/l
RT: 9.524 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 129 Resp: 14007
Ion Ratio Lower Upper
129 100
127 77.7 38.4 115.2









#61
1,2-Dibromoethane
Concen: 18.76 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:107 Resp: 18915
Ion Ratio Lower Upper
107 100
109 93.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.23 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 95 Resp: 59866
Ion Ratio Lower Upper
95 100
174 59.1 0.0 154.6
176 56.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.061 min Scan# 1472
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:117 Resp: 121228
Ion Ratio Lower Upper
117 100
82 61.7 47.4 71.2
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 19.07 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:164 Resp: 16246
Ion Ratio Lower Upper
164 100
166 121.2 103.4 155.2
129 96.4 70.8 106.2
131 99.9 72.5 108.7









#65
Chlorobenzene
Concen: 17.87 ug/l
RT: 10.085 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:112 Resp: 45623
Ion Ratio Lower Upper
112 100
114 32.2 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 18.50 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:131 Resp: 15265
Ion Ratio Lower Upper
131 100
133 91.9 47.7 143.1
119 65.1 32.1 96.5









#67
Ethyl Benzene
Concen: 18.59 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 91 Resp: 87411
Ion Ratio Lower Upper
91 100
106 31.0 24.6 37.0



#68
m/p-Xylenes
Concen: 36.15 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 106 Resp: 63109
Ion Ratio Lower Upper
106 100
91 215.6 162.8 244.2









#69
o-Xylene
Concen: 18.48 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:106 Resp: 31407
Ion Ratio Lower Upper
106 100
91 227.2 108.1 324.2



#70
Styrene
Concen: 18.32 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:104 Resp: 51674
Ion Ratio Lower Upper
104 100
78 57.3 40.0 60.0
103 53.5 43.4 65.0









#71
Bromoform
Concen: 16.74 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:	173	Resp:	7913
Ion Ratio	Lower	Upper	
173	100		
175	44.4	23.9	71.8
254	0.0	0.1	0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:	152	Resp:	51307
Ion Ratio	Lower	Upper	
152	100		
115	79.8	41.1	123.3
150	164.5	0.0	349.0









#73
Isopropylbenzene
Concen: 19.04 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:105 Resp: 82657
Ion Ratio Lower Upper
105 100
120 25.5 13.2 39.6



#74
N-ethyl acetate
Concen: 19.87 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 43 Resp: 45516
Ion Ratio Lower Upper
43 100
70 38.1 39.0 58.6#
55 23.7 21.5 32.3
61 22.2 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 19.08 ug/l
RT: 11.219 min Scan# 1662
Delta R.T. 0.006 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 83 Resp: 28448
Ion Ratio Lower Upper
83 100
131 8.7 5.1 15.2
85 65.7 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 24.35 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 75 Resp: 33004
Ion Ratio Lower Upper
75 100
77 32.9 19.1 57.5









#77
Bromobenzene
Concen: 17.54 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:156 Resp: 15833
Ion Ratio Lower Upper
156 100
77 231.4 82.8 248.6
158 100.4 49.1 147.4



#78
n-propylbenzene
Concen: 19.12 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 91 Resp: 100456
Ion Ratio Lower Upper
91 100
120 20.8 11.5 34.5









#79
2-Chlorotoluene
Concen: 18.87 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 91 Resp: 59840
Ion Ratio Lower Upper
91 100
126 30.0 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 19.49 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:105 Resp: 69465
Ion Ratio Lower Upper
105 100
120 45.7 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 18.85 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion: 75 Resp: 7119
Ion Ratio Lower Upper
75 100
53 133.1 74.5 111.7#
89 44.5 37.9 56.9



#82
4-Chlorotoluene
Concen: 18.80 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 91 Resp: 68913
Ion Ratio Lower Upper
91 100
126 26.4 14.9 44.5









#83
tert-Butylbenzene
Concen: 18.52 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:	119	Resp:	62359
Ion	Ratio	Lower	Upper
119	100		
91	70.0	27.9	83.7
134	23.2	11.3	33.9



#84
1,2,4-Trimethylbenzene
Concen: 19.45 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:	105	Resp:	69797
Ion	Ratio	Lower	Upper
105	100		
120	41.8	22.4	67.3









#85
sec-Butylbenzene
Concen: 18.95 ug/l
RT: 11.896 min Scan# 1773
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:105 Resp: 86024
Ion Ratio Lower Upper
105 100
134 18.5 10.0 29.8



#86
p-Isopropyltoluene
Concen: 18.78 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:119 Resp: 67747
Ion Ratio Lower Upper
119 100
134 25.4 12.9 38.7
91 28.6 11.7 35.0









#87

1,3-Dichlorobenzene
Concen: 18.07 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01




Tgt Ion:146 Resp: 31465
Ion Ratio Lower Upper
146 100
111 49.3 20.2 60.5
148 65.5 31.9 95.7




#88

1,4-Dichlorobenzene
Concen: 17.87 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:146 Resp: 32548
Ion Ratio Lower Upper
146 100
111 47.9 19.6 58.8
148 63.4 31.6 95.0









#89
n-Butylbenzene
Concen: 18.92 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion: 91 Resp: 64895




Ion	Ratio	Lower	Upper
91	100		
92	53.9	27.4	82.0
134	22.0	13.2	39.5




#90
Hexachloroethane
Concen: 16.61 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 117 Resp: 9294

Ion	Ratio	Lower	Upper
117	100		
201	51.0	40.4	121.2









#91
1,2-Dichlorobenzene
Concen: 19.05 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:146 Resp: 32385
Ion Ratio Lower Upper
146 100
111 48.6 20.8 62.2
148 62.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.16 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. -0.001 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion: 75 Resp: 5427
Ion Ratio Lower Upper
75 100
155 62.5 45.3 135.9
157 85.7 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 18.00 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52




Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01


Tgt Ion:180 Resp: 16970
Ion Ratio Lower Upper
180 100
182 96.1 47.7 143.1
145 37.3 15.0 45.0



#94
Hexachlorobutadiene
Concen: 18.10 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:225 Resp: 6686
Ion Ratio Lower Upper
225 100
223 68.4 31.1 93.2
227 59.9 31.6 95.0









#95
Naphthalene
Concen: 18.54 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBSD01

Tgt Ion:128 Resp: 72281

Ion	Ratio	Lower	Upper
128	100		
127	12.1	10.3	15.5
129	11.1	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 18.65 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX023084.D
Acq: 01 Jul 2021 14:52

Tgt Ion:180 Resp: 17714

Ion	Ratio	Lower	Upper
180	100		
182	97.2	47.6	142.9
145	40.3	16.1	48.3

