

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023077.D
 Acq On : 01 Jul 2021 10:59
 Operator : JC/MD
 Sample : VX0701MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701MBL01

Quant Time: Jul 01 14:58:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	56892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115432	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54758	53.036	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.080%
35) Dibromofluoromethane	5.397	113	36460	49.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.480%
50) Toluene-d8	8.653	98	143105	49.813	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.085	95	47935	44.708	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023077.D
Acq On : 01 Jul 2021 10:59
Operator : JC/MD
Sample : VX0701MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0701MBL01

Quant Time: Jul 01 14:58:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

