

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023078.D
 Acq On : 01 Jul 2021 12:29
 Operator : JC/MD
 Sample : VX0701WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701WBL01

Quant Time: Jul 01 14:59:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54995	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	112571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52590	52.693	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.380%
35) Dibromofluoromethane	5.397	113	35085	49.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	137803	49.187	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.085	95	46951	44.903	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.800%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.788	84	1223	1.756	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023078.D
Acq On : 01 Jul 2021 12:29
Operator : JC/MD
Sample : VX0701WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0701WBL01

Quant Time: Jul 01 14:59:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

