

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023085.D
 Acq On : 01 Jul 2021 15:16
 Operator : JC/MD
 Sample : M2882-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRIP-COMP-AUD-1121-1122DL

Quant Time: Jul 01 16:06:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	55682	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	6.769	114	111726	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58994	58.380	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.760%
35) Dibromofluoromethane	5.391	113	36344	51.226	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.460%
50) Toluene-d8	8.653	98	139941	50.327	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.660%
62) 4-Bromofluorobenzene	11.085	95	50231	48.404	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.800%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	5075	24.097	ug/l	# 82
16) Acetone	2.386	43	21084	47.064	ug/l	# 33
25) 2-Butanone	4.568	43	207382	297.686	ug/l	# 89
52) Toluene	8.726	92	3484	1.684	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023085.D
Acq On : 01 Jul 2021 15:16
Operator : JC/MD
Sample : M2882-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRIP-COMP-AUD-1121-1122DL

Quant Time: Jul 01 16:06:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

