

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023088.D
 Acq On : 01 Jul 2021 16:27
 Operator : JC/MD
 Sample : PB137446TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137446TB

Quant Time: Jul 02 04:30:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55185	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	116182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38360	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	61386	61.295	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	122.580%	#
35) Dibromofluoromethane	5.397	113	36617	49.631	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.260%	
50) Toluene-d8	8.653	98	147299	50.942	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.880%	
62) 4-Bromofluorobenzene	11.085	95	54331	50.347	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8205	18.480	ug/l	89
20) Methylene Chloride	2.794	84	8714	10.194	ug/l #	71
25) 2-Butanone	4.574	43	4871	7.055	ug/l #	81
43) Isopropyl Acetate	6.348	43	12464	5.423	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023088.D
Acq On : 01 Jul 2021 16:27
Operator : JC/MD
Sample : PB137446TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137446TB

Quant Time: Jul 02 04:30:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

