

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
 Data File : VX023089.D
 Acq On : 01 Jul 2021 16:50
 Operator : JC/MD
 Sample : PB137454TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137454TB

Quant Time: Jul 02 04:31:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53478	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	110239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59844	61.662	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	123.320%#
35) Dibromofluoromethane	5.404	113	36054	51.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.000%
50) Toluene-d8	8.653	98	140980	51.385	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.760%
62) 4-Bromofluorobenzene	11.085	95	51676	50.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	16000	37.187	ug/l	# 86
20) Methylene Chloride	2.794	84	5297	6.533	ug/l	# 84
43) Isopropyl Acetate	6.355	43	13099	6.007	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070121\
Data File : VX023089.D
Acq On : 01 Jul 2021 16:50
Operator : JC/MD
Sample : PB137454TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137454TB

Quant Time: Jul 02 04:31:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

