

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070121\  
 Data File : VX023099.D  
 Acq On : 01 Jul 2021 20:46  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 IBLK

Quant Time: Jul 02 04:34:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062321W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 23 19:12:26 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 5.562  | 168   | 56728    | 50.000   | ug/l  | # 0.00    |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 120446   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 105056   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 35863    | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 66560    | 64.653   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 129.300%# |
| 35) Dibromofluoromethane    | 5.397  | 113   | 38866    | 50.814   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 101.620%  |
| 50) Toluene-d8              | 8.653  | 98    | 149887   | 50.002   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 100.000%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 54898    | 49.071   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 98.140%   |
| Target Compounds            |        |       |          |          |       |           |
|                             |        |       |          |          |       | Qvalue    |
| 16) Acetone                 | 2.392  | 43    | 1184     | 2.594    | ug/l  | # 68      |
| 20) Methylene Chloride      | 2.794  | 84    | 866      | 1.322    | ug/l  | # 88      |
| 52) Toluene                 | 8.720  | 92    | 557      | 0.250    | ug/l  | 95        |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070121\  
Data File : VX023099.D  
Acq On : 01 Jul 2021 20:46  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
IBLK

Quant Time: Jul 02 04:34:09 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062321W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 23 19:12:26 2021  
Response via : Initial Calibration

