



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070121\  
Data File : VX023075.D  
Acq On : 01 Jul 2021 09:52  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062321W.M  
Title : SW846 8260  
Last Update : Wed Jun 23 19:12:26 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 23.3         | 6184       | PASS                |
| 75             | 95              | 30              | 60              | 58.2         | 15469      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 26563      | PASS                |
| 96             | 95              | 5               | 9               | 6.8          | 1809       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 186        | PASS                |
| 174            | 95              | 50              | 100             | 57.3         | 15212      | PASS                |
| 175            | 174             | 5               | 9               | 6.5          | 995        | PASS                |
| 176            | 174             | 95              | 101             | 98.0         | 14905      | PASS                |
| 177            | 176             | 5               | 9               | 6.6          | 980        | PASS                |