

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
 Data File : VX042102.D
 Acq On : 01 Jul 2024 10:25
 Operator : JC/MD
 Sample : VX0701MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0701MBL01

Quant Time: Jul 01 16:05:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	155010	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	237020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75035	41.979	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.960%
35) Dibromofluoromethane	5.385	113	73735	43.843	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.680%
50) Toluene-d8	8.647	98	283678	48.014	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.079	95	111697	51.286	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
Data File : VX042102.D
Acq On : 01 Jul 2024 10:25
Operator : JC/MD
Sample : VX0701MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0701MBL01

Quant Time: Jul 01 16:05:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

