

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
 Data File : VX042107.D
 Acq On : 01 Jul 2024 12:25
 Operator : JC/MD
 Sample : P3127-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-EB-20240625RE

Quant Time: Jul 01 16:06:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148402	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	227157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73711	43.074	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.140%
35) Dibromofluoromethane	5.391	113	72164	44.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.647	98	264806	46.766	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.540%
62) 4-Bromofluorobenzene	11.079	95	98483	47.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.360%
Target Compounds						
16) Acetone	2.386	43	4227	5.625	ug/l	90
44) Trichloroethene	7.129	130	325	0.210	ug/l	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
Data File : VX042107.D
Acq On : 01 Jul 2024 12:25
Operator : JC/MD
Sample : P3127-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-EB-20240625RE

Quant Time: Jul 01 16:06:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

