

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
 Data File : VX042109.D
 Acq On : 01 Jul 2024 13:12
 Operator : JC/MD
 Sample : P3092-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20240626

Quant Time: Jul 02 02:04:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	166673	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	253481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81628	42.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.940%
35) Dibromofluoromethane	5.385	113	79500	44.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.400%
50) Toluene-d8	8.647	98	295509	46.769	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.540%
62) 4-Bromofluorobenzene	11.079	95	108726	46.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.360%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3842	4.553	ug/l	96
20) Methylene Chloride	2.788	84	1779	0.963	ug/l #	86
95) Naphthalene	13.774	128	18002	2.790	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
Data File : VX042109.D
Acq On : 01 Jul 2024 13:12
Operator : JC/MD
Sample : P3092-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20240626

Quant Time: Jul 02 02:04:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

