

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
 Data File : VX042120.D
 Acq On : 01 Jul 2024 17:29
 Operator : JC/MD
 Sample : P3092-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW4-20240626

Quant Time: Jul 02 02:09:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136390	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	210336	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67928	43.191	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.380%
35) Dibromofluoromethane	5.385	113	65981	44.210	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.420%
50) Toluene-d8	8.647	98	243659	46.473	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.940%
62) 4-Bromofluorobenzene	11.079	95	94085	48.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.360%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2403	3.480	ug/l	98
20) Methylene Chloride	2.788	84	645	0.427	ug/l #	89
40) Benzene	6.038	78	2291	0.429	ug/l	93
67) Ethyl Benzene	10.195	91	1725	0.254	ug/l #	78
68) m/p-Xylenes	10.305	106	1460	0.533	ug/l	91
69) o-Xylene	10.640	106	685	0.253	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	2236	0.408	ug/l	93
95) Naphthalene	13.780	128	3249	1.162	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
Data File : VX042120.D
Acq On : 01 Jul 2024 17:29
Operator : JC/MD
Sample : P3092-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW4-20240626

Quant Time: Jul 02 02:09:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

