

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
 Data File : VX042122.D
 Acq On : 01 Jul 2024 18:15
 Operator : JC/MD
 Sample : P3106-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-1

Quant Time: Jul 02 02:10:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	134887	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	209393	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67273	43.251	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.500%
35) Dibromofluoromethane	5.385	113	65035	43.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.540%
50) Toluene-d8	8.646	98	243400	46.632	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.079	95	92725	48.193	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.380%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1473	2.157	ug/l	# 88
18) Methyl Acetate	2.715	43	2047	1.099	ug/l	96
20) Methylene Chloride	2.788	84	1490	0.996	ug/l	# 75
95) Naphthalene	13.780	128	4617	1.364	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070124\
Data File : VX042122.D
Acq On : 01 Jul 2024 18:15
Operator : JC/MD
Sample : P3106-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-1

Quant Time: Jul 02 02:10:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

