

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
 Data File : VX046849.D
 Acq On : 01 Jul 2025 11:03
 Operator : JC/MD
 Sample : Q2461-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-3

Quant Time: Jul 02 01:39:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	224959	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	88565	47.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.360%
35) Dibromofluoromethane	5.397	113	73090	49.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.220%
50) Toluene-d8	8.647	98	262232	49.038	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.079	95	102796	51.569	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	99079	159.968	ug/l	98
25) 2-Butanone	4.580	43	15504	17.699	ug/l	94
51) 4-Methyl-2-Pentanone	8.580	43	8772	4.739	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
Data File : VX046849.D
Acq On : 01 Jul 2025 11:03
Operator : JC/MD
Sample : Q2461-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-3

Quant Time: Jul 02 01:39:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

