

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
 Data File : VX046852.D
 Acq On : 01 Jul 2025 12:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 02 01:40:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	122077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	211259	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	99700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	93070	55.119	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.240%
35) Dibromofluoromethane	5.397	113	74839	53.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.080%
50) Toluene-d8	8.647	98	254800	50.739	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.480%
62) 4-Bromofluorobenzene	11.079	95	100654	53.769	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
Data File : VX046852.D
Acq On : 01 Jul 2025 12:15
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 02 01:40:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

