

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
 Data File : VX046856.D
 Acq On : 01 Jul 2025 13:39
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 02 01:41:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	238505	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	111484	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96685	48.702	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.400%
35) Dibromofluoromethane	5.397	113	79511	50.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.780%
50) Toluene-d8	8.653	98	286461	50.527	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.060%
62) 4-Bromofluorobenzene	11.079	95	110112	52.102	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070125\
Data File : VX046856.D
Acq On : 01 Jul 2025 13:39
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 02 01:41:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

