

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070218\
 Data File : VX003124.D
 Acq On : 02 Jul 2018 13:49
 Operator : JC/MD
 Sample : J3738-04DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S22-0041DL

Quant Time: Jul 03 09:04:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	284461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	396999	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	366023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	218143	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180807	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
35) Dibromofluoromethane	5.51	113	149617	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
50) Toluene-d8	8.72	98	553264	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	11.14	95	200569	45.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.60%	
Target Compounds						
39) Methylcyclohexane	7.46	83	11029	2.17	ug/l	97
67) Ethyl Benzene	10.26	91	692343	45.08	ug/l	100
68) m/p-Xylenes	10.37	106	200081	33.80	ug/l	98
73) Isopropylbenzene	11.02	105	129448	8.82	ug/l	99
78) n-propylbenzene	11.36	91	123145	7.37	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	138093	10.78	ug/l	98
95) Naphthalene	13.84	128	133218	8.92	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070218\
Data File : VX003124.D
Acq On : 02 Jul 2018 13:49
Operator : JC/MD
Sample : J3738-04DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S22-0041DL

Quant Time: Jul 03 09:04:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

