

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070218\
 Data File : VX003129.D
 Acq On : 02 Jul 2018 16:10
 Operator : JC/MD
 Sample : J3772-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP18-GMZ5-ER1

Quant Time: Jul 13 12:24:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	365960	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186711	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168038	45.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.30%	
35) Dibromofluoromethane	5.51	113	141037	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	8.72	98	505185	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	11.14	95	175689	43.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.10%	
Target Compounds						
16) Acetone	2.45	43	26827	17.41	ug/l	98
25) 2-Butanone	4.72	43	15301	6.69	ug/l	97
29) Tetrahydrofuran	5.18	42	48082	32.78	ug/l	95
37) Ethyl Acetate	4.87	43	3322	0.77	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070218\
Data File : VX003129.D
Acq On : 02 Jul 2018 16:10
Operator : JC/MD
Sample : J3772-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP18-GMZ5-ER1

Quant Time: Jul 13 12:24:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

