

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070218\
 Data File : VX003132.D
 Acq On : 02 Jul 2018 17:30
 Operator : JC/MD
 Sample : J3772-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP18-CHY02

Quant Time: Jul 03 11:12:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	367704	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	215997	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	168999	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
35) Dibromofluoromethane	5.51	113	141561	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
50) Toluene-d8	8.72	98	518195	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.14	95	197311	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	94054	61.68	ug/l	95
17) Carbon Disulfide	2.57	76	12762	1.82	ug/l	95
25) 2-Butanone	4.71	43	108723	48.05	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	8851	2.76	ug/l	54
40) Benzene	6.15	78	11989	1.02	ug/l	95
44) Trichloroethene	7.22	130	10157	2.91	ug/l	92
59) 2-Hexanone	9.50	43	129321	37.35	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070218\
Data File : VX003132.D
Acq On : 02 Jul 2018 17:30
Operator : JC/MD
Sample : J3772-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP18-CHY02

Quant Time: Jul 03 11:12:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

