

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070218\
 Data File : VX003134.D
 Acq On : 02 Jul 2018 18:24
 Operator : JC/MD
 Sample : J3772-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP18-CHY05

Quant Time: Jul 03 11:14:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	263567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	364659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	346068	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213547	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168586	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
35) Dibromofluoromethane	5.51	113	140207	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	8.72	98	514687	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.14	95	194872	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
16) Acetone	2.45	43	51473	34.12	ug/l	97
17) Carbon Disulfide	2.57	76	6361	0.92	ug/l	97
25) 2-Butanone	4.73	43	29949	13.38	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	42754	13.47	ug/l	96
40) Benzene	6.15	78	9950	0.86	ug/l	94
44) Trichloroethene	7.22	130	15572	4.50	ug/l	98
59) 2-Hexanone	9.50	43	56717	16.52	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070218\
Data File : VX003134.D
Acq On : 02 Jul 2018 18:24
Operator : JC/MD
Sample : J3772-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP18-CHY05

Quant Time: Jul 03 11:14:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

