

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070218\
 Data File : VX003120.D
 Acq On : 02 Jul 2018 11:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 07:03:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.497	0.515	-3.6	125	0.00
3 P	Chloromethane	0.521	0.517	0.8	118	0.00
4 C	Vinyl Chloride	0.565	0.577	-2.1#	120	0.00
5 T	Bromomethane	0.349	0.168	51.9#	71	0.00
6 T	Chloroethane	0.376	0.372	1.1	117	0.00
7 T	Trichlorofluoromethane	0.954	0.913	4.3	106	0.00
8 T	Diethyl Ether	0.347	0.346	0.3	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.546	1.3	113	0.00
10 T	Methyl Iodide	0.515	0.460	10.7	99	0.00
11 T	Tert butyl alcohol	0.138	0.125	9.4	99	0.00
12 CM	1,1-Dichloroethene	0.493	0.484	1.8#	112	0.00
13 T	Acrolein	0.082	0.059	28.0#	87	0.00
14 T	Allyl chloride	0.972	0.954	1.9	109	0.00
15 T	Acrylonitrile	0.294	0.285	3.1	106	0.00
16 T	Acetone	0.286	0.378	-32.2#	145	0.00
17 T	Carbon Disulfide	1.314	1.402	-6.7	124	0.00
18 T	Methyl Acetate	0.800	0.662	17.3	86	0.00
19 T	Methyl tert-butyl Ether	1.773	1.730	2.4	107	0.00
20 T	Methylene Chloride	0.557	0.543	2.5	110	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.526	2.0	112	0.00
22 T	Diisopropyl ether	1.858	1.910	-2.8	113	0.00
23 T	Vinyl Acetate	1.573	1.695	-7.8	119	0.00
24 P	1,1-Dichloroethane	1.010	0.969	4.1	107	0.00
25 T	2-Butanone	0.425	0.492	-15.8	126	0.00
26 T	2,2-Dichloropropane	0.821	0.860	-4.8	115	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.624	-3.7	116	0.00
28 T	Bromochloromethane	0.483	0.466	3.5	102	0.00
29 T	Tetrahydrofuran	0.272	0.269	1.1	109	0.00
30 C	Chloroform	1.036	1.022	1.4#	109	0.00
31 T	Cyclohexane	0.867	0.891	-2.8	114	0.00
32 T	1,1,1-Trichloroethane	0.912	0.921	-1.0	110	0.00
33 S	1,2-Dichloroethane-d4	0.691	0.641	7.2	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.396	0.406	-2.5	106	0.00
36 T	1,1-Dichloropropene	0.548	0.588	-7.3	114	0.00
37 T	Ethyl Acetate	0.590	0.650	-10.2	115	0.00
38 T	Carbon Tetrachloride	0.602	0.642	-6.6	111	0.00
39 T	Methylcyclohexane	0.639	0.699	-9.4	114	0.00
40 TM	Benzene	1.593	1.703	-6.9	111	0.00
41 T	Methacrylonitrile	0.340	0.361	-6.2	109	0.00
42 TM	1,2-Dichloroethane	0.644	0.640	0.6	105	0.00
43 T	Isopropyl Acetate	0.990	1.031	-4.1	109	0.00
44 TM	Trichloroethene	0.475	0.507	-6.7	114	0.00
45 C	1,2-Dichloropropane	0.423	0.445	-5.2#	114	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX070218\
 Data File : VX003120.D
 Acq On : 02 Jul 2018 11:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 07:03:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.293	0.305	-4.1	108	0.00
47 T	Bromodichloromethane	0.542	0.584	-7.7	109	0.00
48 T	Methyl methacrylate	0.510	0.532	-4.3	105	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	106	0.00
50 S	Toluene-d8	1.446	1.471	-1.7	104	0.00
51 T	4-Methyl-2-Pentanone	0.619	0.652	-5.3	106	0.00
52 CM	Toluene	1.021	1.107	-8.4#	111	0.00
53 T	t-1,3-Dichloropropene	0.614	0.700	-14.0	112	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.653	-10.5	110	0.00
55 T	1,1,2-Trichloroethane	0.419	0.451	-7.6	111	0.00
56 T	Ethyl methacrylate	0.613	0.686	-11.9	113	0.00
57 T	1,3-Dichloropropane	0.689	0.728	-5.7	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.296	0.350	-18.2	117	0.00
59 T	2-Hexanone	0.471	0.530	-12.5	112	0.00
60 T	Dibromochloromethane	0.441	0.497	-12.7	110	0.00
61 T	1,2-Dibromoethane	0.440	0.484	-10.0	112	0.00
62 S	4-Bromofluorobenzene	0.558	0.576	-3.2	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.556	0.601	-8.1	112	0.00
65 PM	Chlorobenzene	1.270	1.338	-5.4	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.461	0.502	-8.9	110	0.00
67 C	Ethyl Benzene	2.098	2.290	-9.2#	110	0.00
68 T	m/p-Xylenes	0.809	0.898	-11.0	111	0.00
69 T	o-Xylene	0.787	0.874	-11.1	111	0.00
70 T	Styrene	1.279	1.456	-13.8	112	0.00
71 P	Bromoform	0.363	0.421	-16.0	113	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.364	3.603	-7.1	110	0.00
74 T	N-amyl acetate	1.646	1.655	-0.5	109	0.00
75 P	1,1,2,2-Tetrachloroethane	1.036	1.079	-4.2	112	0.00
76 T	1,2,3-Trichloropropane	0.906	1.072	-18.3	124	0.00
77 T	Bromobenzene	0.958	1.003	-4.7	111	0.00
78 T	n-propylbenzene	3.829	4.113	-7.4	109	0.00
79 T	2-Chlorotoluene	2.318	2.416	-4.2	109	0.00
80 T	1,3,5-Trimethylbenzene	2.833	3.065	-8.2	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.267	0.304	-13.9	109	0.00
82 T	4-Chlorotoluene	2.744	2.894	-5.5	109	0.00
83 T	tert-Butylbenzene	2.792	2.985	-6.9	109	0.00
84 T	1,2,4-Trimethylbenzene	2.935	3.152	-7.4	110	0.00
85 T	sec-Butylbenzene	3.387	3.567	-5.3	107	0.00
86 T	p-Isopropyltoluene	3.045	3.303	-8.5	108	0.00
87 T	1,3-Dichlorobenzene	1.749	1.854	-6.0	111	0.00
88 T	1,4-Dichlorobenzene	1.800	1.863	-3.5	111	0.00
89 T	n-Butylbenzene	2.650	2.830	-6.8	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070218\
Data File : VX003120.D
Acq On : 02 Jul 2018 11:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 03 07:03:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.453	0.456	-0.7	101	0.00
91 T	1,2-Dichlorobenzene	1.766	1.848	-4.6	110	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.234	-0.4	101	0.00
93 T	1,2,4-Trichlorobenzene	1.285	1.391	-8.2	112	0.00
94 T	Hexachlorobutadiene	0.641	0.692	-8.0	111	0.00
95 T	Naphthalene	3.423	3.806	-11.2	111	0.00
96 T	1,2,3-Trichlorobenzene	1.284	1.389	-8.2	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6