

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070218\
 Data File : VX003126.D
 Acq On : 02 Jul 2018 14:42
 Operator : JC/MD
 Sample : J3787-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP18-GMZ5-TB2

Quant Time: Jul 03 11:05:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	275216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	379687	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349082	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195575	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174302	45.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.64%	
35) Dibromofluoromethane	5.51	113	143716	47.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.60%	
50) Toluene-d8	8.72	98	528856	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
62) 4-Bromofluorobenzene	11.14	95	185132	43.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.44%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070218\
Data File : VX003126.D
Acq On : 02 Jul 2018 14:42
Operator : JC/MD
Sample : J3787-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP18-GMZ5-TB2

Quant Time: Jul 03 11:05:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.00 min

Lab File: VX003126.D

Acq: 02 Jul 2018 14:42

Instrument :

MSVOA_X

ClientSampleId :


SP18-GMZ5-TB2

Tot Ion:168 Resp: 275216

Ion Ratio Lower Upper

168 100

99 49.6 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 45.82 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX003126.D


Acq: 02 Jul 2018 14:42

Tgt Ion: 65 Resp: 174302

Ion Ratio Lower Upper

65 100

67 52.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000


60000


40000

20000

0

Time--> 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX003126.D

Acq: 02 Jul 2018 14:42

Instrument :

MSVOA_X

ClientSampleId :

SP18-GMZ5-TB2


Tot Ion:114 Resp: 379687

Ion Ratio Lower Upper

114 100

63 19.5 0.0 41.4




88 14.6 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.80 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX003126.D

Acq: 02 Jul 2018 14:42


Tgt Ion:113 Resp: 143716

Ion Ratio Lower Upper

113 100

111 103.0 81.4 122.0



192 25.3 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.17 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003126.D
Acq: 02 Jul 2018 14:42




Instrument :
MSVOA_X
ClientSampleId :
SP18-GMZ5-TB2


Tot Ion: 98 Resp: 528856
Ion Ratio Lower Upper
98 100
100 64.0 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 43.72 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003126.D
Acq: 02 Jul 2018 14:42

Tgt Ion: 95 Resp: 185132
Ion Ratio Lower Upper
95 100
174 98.2 0.0 187.4
176 94.3 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX003126.D
Acq: 02 Jul 2018 14:42

Instrument :
MSVOA_X
ClientSampleId :
SP18-GMZ5-TB2

Tot Ion:117 Resp: 349082
Ion Ratio Lower Upper
117 100
82 53.6 45.0 67.4
119 31.9 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.01 min
Lab File: VX003126.D
Acq: 02 Jul 2018 14:42

Tgt Ion:152 Resp: 195575
Ion Ratio Lower Upper
152 100
115 54.7 40.1 120.2
150 157.9 0.0 347.2

