

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070219\
 Data File : VX010616.D
 Acq On : 02 Jul 2019 14:50
 Operator : JC/SP
 Sample : K3638-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP

Quant Time: Jul 03 03:21:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	369147	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	334596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153446	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118062	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
35) Dibromofluoromethane	5.50	113	112738	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
50) Toluene-d8	8.71	98	435858	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	11.13	95	144100	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
16) Acetone	2.44	43	3444	3.068	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070219\
Data File : VX010616.D
Acq On : 02 Jul 2019 14:50
Operator : JC/SP
Sample : K3638-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP

Quant Time: Jul 03 03:21:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

