

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070219\
 Data File : VX010625.D
 Acq On : 02 Jul 2019 18:21
 Operator : JC/SP
 Sample : K3636-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP19-MMW0018B

Quant Time: Jul 03 03:58:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	234088	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	366574	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	329895	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151303	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116187	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
35) Dibromofluoromethane	5.50	113	110116	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	8.71	98	429010	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	11.13	95	142227	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
Target Compounds						
16) Acetone	2.44	43	2903	2.639	ug/l #	Qvalue 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070219\
Data File : VX010625.D
Acq On : 02 Jul 2019 18:21
Operator : JC/SP
Sample : K3636-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP19-MMW0018B

Quant Time: Jul 03 03:58:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

