

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070219\
 Data File : VX010627.D
 Acq On : 02 Jul 2019 19:07
 Operator : JC/SP
 Sample : PB121086TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121086TB

Quant Time: Jul 03 04:05:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	372885	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	343339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153660	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	119629	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
35) Dibromofluoromethane	5.50	113	114501	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
50) Toluene-d8	8.71	98	446071	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.13	95	149143	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.04	59	5860	10.067	ug/l	95
16) Acetone	2.44	43	31690	28.426	ug/l	99
20) Methylene Chloride	2.85	84	5555	2.421	ug/l	91
25) 2-Butanone	4.68	43	8622	5.272	ug/l	98
29) Tetrahydrofuran	5.13	42	41411	40.882	ug/l	97
43) Isopropyl Acetate	6.45	43	14390	2.853	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	8720	1.015	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070219\
Data File : VX010627.D
Acq On : 02 Jul 2019 19:07
Operator : JC/SP
Sample : PB121086TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121086TB

Quant Time: Jul 03 04:05:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

