

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070220\
 Data File : VX017151.D
 Acq On : 02 Jul 2020 16:17
 Operator : JC/SP
 Sample : PB129937TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129937TB

Quant Time: Jul 03 01:47:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	220430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	363567	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	372878	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	141775	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
35) Dibromofluoromethane	5.47	113	122087	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
50) Toluene-d8	8.70	98	445796	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
62) 4-Bromofluorobenzene	11.12	95	183237	55.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.24%	
Target Compounds						
16) Acetone	2.42	43	5436	5.347	ug/l	94
20) Methylene Chloride	2.84	84	5478	1.592	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070220\
Data File : VX017151.D
Acq On : 02 Jul 2020 16:17
Operator : JC/SP
Sample : PB129937TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129937TB

Quant Time: Jul 03 01:47:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

