

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070220\
 Data File : VX017152.D
 Acq On : 02 Jul 2020 16:40
 Operator : JC/SP
 Sample : L3192-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW180-SOIL-600-604

Quant Time: Jul 03 01:50:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	223178	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	367684	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	378863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	201006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	144023	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
35) Dibromofluoromethane	5.47	113	124562	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
50) Toluene-d8	8.70	98	451451	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	11.12	95	187450	55.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.50%	
Target Compounds						
16) Acetone	2.42	43	9265	9.002	ug/l	95
20) Methylene Chloride	2.84	84	17341	6.668	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070220\
Data File : VX017152.D
Acq On : 02 Jul 2020 16:40
Operator : JC/SP
Sample : L3192-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW180-SOIL-600-604

Quant Time: Jul 03 01:50:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

