

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070220\
 Data File : VX017155.D
 Acq On : 02 Jul 2020 17:48
 Operator : JC/SP
 Sample : L3192-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW180-SOIL-614-616

Quant Time: Jul 03 01:59:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	206015	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	344462	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	358619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	190561	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	136471	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
35) Dibromofluoromethane	5.46	113	116198	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
50) Toluene-d8	8.70	98	422674	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
62) 4-Bromofluorobenzene	11.12	95	174382	55.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.72%	
Target Compounds						
16) Acetone	2.42	43	11746	12.363	ug/l	91
20) Methylene Chloride	2.84	84	33541	14.842	ug/l	94
79) 2-Chlorotoluene	11.40	91	15798	1.849	ug/l	99
82) 4-Chlorotoluene	11.49	91	16482	1.614	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070220\
Data File : VX017155.D
Acq On : 02 Jul 2020 17:48
Operator : JC/SP
Sample : L3192-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW180-SOIL-614-616

Quant Time: Jul 03 01:59:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

