

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023108.D
 Acq On : 02 Jul 2021 13:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 05 02:52:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	232605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	377951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	339719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155491	49.738	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.480%
35) Dibromofluoromethane	5.397	113	120541	49.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.653	98	454412	49.912	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.820%
62) 4-Bromofluorobenzene	11.085	95	162468	46.225	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.440%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	1069	0.507	ug/l	87
3) Chloromethane	1.300	50	1072	0.514	ug/l	89
4) Vinyl Chloride	1.374	62	850	0.385	ug/l #	59
5) Bromomethane	1.605	94	1612	Below Cal	#	60
6) Chloroethane	1.697	64	458	0.317	ug/l #	51
7) Trichlorofluoromethane	1.892	101	1321	0.344	ug/l	90
9) 1,1,2-Trichlorotrifluo...	2.337	101	1339	0.596	ug/l	93
10) Methyl Iodide	2.465	142	2329	0.844	ug/l	94
11) Tert butyl alcohol	2.953	59	3037	4.618	ug/l #	81
12) 1,1-Dichloroethene	2.325	96	1212	0.585	ug/l	90
13) Acrolein	2.245	56	981	2.240	ug/l #	71
14) Allyl chloride	2.672	41	1246	0.346	ug/l #	78
15) Acrylonitrile	3.075	53	1983	1.551	ug/l #	87
16) Acetone	2.392	43	1940	1.569	ug/l	90
17) Carbon Disulfide	2.514	76	8611	1.969	ug/l	97
18) Methyl Acetate	2.715	43	785	0.261	ug/l #	70
20) Methylene Chloride	2.794	84	2398	0.901	ug/l #	68
21) trans-1,2-Dichloroethene	3.105	96	1674	0.748	ug/l	88
23) Vinyl Acetate	3.739	43	3269	0.508	ug/l	99
25) 2-Butanone	4.593	43	1990	1.067	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	1086	0.400	ug/l	76
28) Bromochloromethane	4.903	49	675	0.331	ug/l #	93
29) Tetrahydrofuran	5.050	42	1473	1.229	ug/l #	81
31) Cyclohexane	5.470	56	1578	0.422	ug/l #	82
36) 1,1-Dichloropropene	5.714	75	1979	0.587	ug/l #	41
38) Carbon Tetrachloride	5.690	117	806	0.229	ug/l #	74
39) Methylcyclohexane	7.391	83	1885	0.478	ug/l	91
40) Benzene	6.050	78	2161	0.221	ug/l #	87
42) 1,2-Dichloroethane	6.111	62	1421	0.362	ug/l #	73
44) Trichloroethene	7.135	130	1398	0.504	ug/l	91
46) Dibromomethane	7.592	93	866	0.475	ug/l #	83
49) 1,4-Dioxane	7.671	88	331	4.903	ug/l #	86
51) 4-Methyl-2-Pentanone	8.592	43	1303	0.344	ug/l	93
52) Toluene	8.726	92	1667	0.263	ug/l	98
53) t-1,3-Dichloropropene	8.988	75	1045	2.971	ug/l	99
54) cis-1,3-Dichloropropene	8.378	75	1028	2.397	ug/l #	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023108.D
 Acq On : 02 Jul 2021 13:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 05 02:52:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
58) 2-Chloroethyl Vinyl ether	8.250	63	1159	5.198	ug/l	95
59) 2-Hexanone	9.439	43	1751	0.602	ug/l	73
60) Dibromochloromethane	9.531	129	244	3.190	ug/l	96
61) 1,2-Dibromoethane	9.610	107	563	0.206	ug/l	94
64) Tetrachloroethene	9.275	164	1348	0.497	ug/l #	70
65) Chlorobenzene	10.085	112	2614	0.389	ug/l	92
67) Ethyl Benzene	10.201	91	4304	0.363	ug/l	95
68) m/p-Xylenes	10.311	106	3301	0.728	ug/l	96
69) o-Xylene	10.646	106	1311	0.287	ug/l	72
70) Styrene	10.665	104	2305	0.309	ug/l	95
71) Bromoform	10.799	173	140	4.119	ug/l #	29
73) Isopropylbenzene	10.969	105	4060	0.392	ug/l	96
74) N-amyl acetate	10.848	43	927	0.211	ug/l #	73
76) 1,2,3-Trichloropropane	11.085	75	86567	29.245	ug/l #	39
77) Bromobenzene	11.207	156	1236	0.478	ug/l	93
78) n-propylbenzene	11.305	91	6335	0.529	ug/l	97
79) 2-Chlorotoluene	11.366	91	3327	0.449	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	3914	0.439	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.085	75	86567	78.644	ug/l #	12
82) 4-Chlorotoluene	11.457	91	5429	0.649	ug/l	100
83) tert-Butylbenzene	11.719	119	3946	0.466	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	3959	0.449	ug/l	99
85) sec-Butylbenzene	11.896	105	5985	0.549	ug/l	94
86) p-Isopropyltoluene	12.012	119	5350	0.585	ug/l #	86
87) 1,3-Dichlorobenzene	11.975	146	3344	0.688	ug/l	94
88) 1,4-Dichlorobenzene	11.975	146	3344	0.684	ug/l	92
89) n-Butylbenzene	12.335	91	6334	0.791	ug/l	95
90) Hexachloroethane	12.542	117	335	3.197	ug/l	74
91) 1,2-Dichlorobenzene	12.341	146	2608	0.540	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.951	75	147	2.743	ug/l	62
93) 1,2,4-Trichlorobenzene	13.591	180	4099	1.377	ug/l	91
94) Hexachlorobutadiene	13.725	225	2394	1.899	ug/l	94
95) Naphthalene	13.780	128	9084	2.082	ug/l #	95
96) 1,2,3-Trichlorobenzene	13.963	180	3480	1.162	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023108.D
 Acq On : 02 Jul 2021 13:49
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 05 02:52:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 232605
Ion Ratio Lower Upper
168 100
99 62.7 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 0.507 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 85 Resp: 1069
Ion Ratio Lower Upper
85 100
87 25.6 16.4 49.2









#3
Chloromethane
Concen: 0.514 ug/l
RT: 1.300 min Scan# 35
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 50 Resp: 1072
Ion Ratio Lower Upper
50 100
52 39.3 26.4 39.6



#4
Vinyl Chloride
Concen: 0.385 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 62 Resp: 850
Ion Ratio Lower Upper
62 100
64 54.9 25.7 38.5#









#5
Bromomethane
Concen: Below Cal
RT: 1.605 min Scan# 85
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 94 Resp: 1612
Ion Ratio Lower Upper
94 100
96 56.7 76.4 114.6#



#6
Chloroethane
Concen: 0.317 ug/l
RT: 1.697 min Scan# 100
Delta R.T. 0.018 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 64 Resp: 458
Ion Ratio Lower Upper
64 100
66 59.1 25.4 38.0#









#7
Trichlorofluoromethane
Concen: 0.344 ug/l
RT: 1.892 min Scan# 132
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:101 Resp: 1321
Ion Ratio Lower Upper
101 100
103 74.1 53.2 79.8



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.596 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49


Tgt Ion:101 Resp: 1339
Ion Ratio Lower Upper
101 100
85 44.1 34.9 52.3
151 67.8 61.4 92.2








#10
Methyl Iodide
Concen: 0.844 ug/l
RT: 2.465 min Scan# 226
Delta R.T. 0.012 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Instrument :
MSVOA_X
ClientSampled :
IBLK







Tgt Ion:142 Resp: 2329
Ion Ratio Lower Upper
142 100
127 45.6 33.3 49.9
141 12.2 11.4 17.0



#11
Tert butyl alcohol
Concen: 4.618 ug/l
RT: 2.953 min Scan# 306
Delta R.T. -0.030 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 59 Resp: 3037
Ion Ratio Lower Upper
59 100
57 2.7 7.8 11.8#





#12

1,1-Dichloroethene

Concen: 0.585 ug/l

RT: 2.325 min Scan# 203





Delta R.T. 0.006 min

Lab File: VX023108.D

Acq: 02 Jul 2021 13:49

Tgt Ion: 96 Resp: 1212

Ion	Ratio	Lower	Upper
96	100		
61	138.9	124.6	187.0
98	61.2	50.6	76.0



#13

Acrolein

Concen: 2.240 ug/l

RT: 2.245 min Scan# 190




Delta R.T. 0.000 min


Lab File: VX023108.D

Acq: 02 Jul 2021 13:49

Tgt Ion: 56 Resp: 981

Ion	Ratio	Lower	Upper
56	100		
55	94.7	56.4	84.6#









#14
 Allyl chloride
 Concen: 0.346 ug/l
 RT: 2.672 min Scan# 260
 Delta R.T. 0.006 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion: 41	Resp: 1246
Ion Ratio	Lower Upper
41 100	
39 78.7	51.6 77.4#
76 21.5	30.5 45.7#



#15
 Acrylonitrile
 Concen: 1.551 ug/l
 RT: 3.075 min Scan# 326
 Delta R.T. -0.000 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49

Tgt Ion: 53	Resp: 1983
Ion Ratio	Lower Upper
53 100	
52 91.3	65.3 97.9
51 46.1	28.2 42.2#









#16
Acetone
Concen: 1.569 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 1940
Ion Ratio Lower Upper
43 100
58 40.6 27.8 41.8



#17
Carbon Disulfide
Concen: 1.969 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 76 Resp: 8611
Ion Ratio Lower Upper
76 100
78 8.0 7.2 10.8









#18
Methyl Acetate
Concen: 0.261 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 785
Ion Ratio Lower Upper
43 100
74 11.6 21.9 32.9#



#20
Methylene Chloride
Concen: 0.901 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 84 Resp: 2398
Ion Ratio Lower Upper
84 100
49 150.4 88.6 132.8#
51 43.2 27.0 40.6#
86 85.5 50.3 75.5#









#21
trans-1,2-Dichloroethene
Concen: 0.748 ug/l
RT: 3.105 min Scan# 331
Delta R.T. 0.012 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 96 Resp: 1674
Ion Ratio Lower Upper
96 100
61 148.6 107.7 161.5
98 53.0 50.4 75.6



#23
Vinyl Acetate
Concen: 0.508 ug/l
RT: 3.739 min Scan# 435
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 43 Resp: 3269
Ion Ratio Lower Upper
43 100
86 12.8 9.9 14.9









#25
2-Butanone
Concen: 1.067 ug/l
RT: 4.593 min Scan# 575
Delta R.T. 0.025 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 1990
Ion Ratio Lower Upper
43 100
72 26.7 23.7 35.5



#27
cis-1,2-Dichloroethene
Concen: 0.400 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 96 Resp: 1086
Ion Ratio Lower Upper
96 100
61 182.5 0.0 285.6
98 68.8 0.0 129.6









#28
Bromochloromethane
Concen: 0.331 ug/l
RT: 4.903 min Scan# 626
Delta R.T. -0.001 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 49 Resp: 675
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.0
130 78.8 68.2 102.2



#29
Tetrahydrofuran
Concen: 1.229 ug/l
RT: 5.050 min Scan# 650
Delta R.T. 0.031 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 42 Resp: 1473
Ion Ratio Lower Upper
42 100
72 39.2 41.7 62.5#
71 34.2 38.4 57.6#









#31
Cyclohexane
Concen: 0.422 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.007 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 56 Resp: 1578
Ion Ratio Lower Upper
56 100
69 28.3 25.3 37.9
84 68.9 71.4 107.0#



#33
1,2-Dichloroethane-d4
Concen: 49.738 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 65 Resp: 155491
Ion Ratio Lower Upper
65 100
67 51.0 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:114 Resp: 377951
Ion Ratio Lower Upper
114 100
63 20.4 0.0 40.8
88 17.1 0.0 32.8



#35
Dibromofluoromethane
Concen: 49.389 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:113 Resp: 120541
Ion Ratio Lower Upper
113 100
111 100.6 81.9 122.9
192 18.7 16.7 25.1









#36
1,1-Dichloropropene
Concen: 0.587 ug/l
RT: 5.714 min Scan# 759
Delta R.T. 0.012 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 75 Resp: 1979
Ion Ratio Lower Upper
75 100
110 0.0 18.1 54.4#
77 0.0 25.3 37.9#



#38
Carbon Tetrachloride
Concen: 0.229 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 117 Resp: 806
Ion Ratio Lower Upper
117 100
119 67.9 76.2 114.2#
121 19.5 24.8 37.2#









#39
Methylcyclohexane
Concen: 0.478 ug/l
RT: 7.391 min Scan# 1034
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 83 Resp: 1885
Ion Ratio Lower Upper
83 100
55 82.1 58.9 88.3
98 51.9 38.1 57.1



#40
Benzene
Concen: 0.221 ug/l
RT: 6.050 min Scan# 814
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 78 Resp: 2161
Ion Ratio Lower Upper
78 100
77 17.2 18.9 28.3#





#42

1,2-Dichloroethane

Concen: 0.362 ug/l

RT: 6.111 min Scan# 824

Delta R.T. 0.013 min

Lab File: VX023108.D





Acq: 02 Jul 2021 13:49

Tgt Ion: 62 Resp: 1421

Ion Ratio Lower Upper

62 100

98 0.0 0.0 20.2



#44

Trichloroethene

Concen: 0.504 ug/l

RT: 7.135 min Scan# 992

Delta R.T. -0.000 min

Lab File: VX023108.D




Acq: 02 Jul 2021 13:49


Tgt Ion: 130 Resp: 1398

Ion Ratio Lower Upper

130 100

95 90.3 0.0 198.0









#46
Dibromomethane
Concen: 0.475 ug/l
RT: 7.592 min Scan# 1067
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 93 Resp: 866
Ion Ratio Lower Upper
93 100
95 83.3 66.8 100.2
174 72.7 83.1 124.7#



#49
1,4-Dioxane
Concen: 4.903 ug/l
RT: 7.671 min Scan# 1080
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 88 Resp: 331
Ion Ratio Lower Upper
88 100
43 43.5 23.2 34.8#
58 76.1 55.0 82.4









#50
Toluene-d8
Concen: 49.912 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 98 Resp: 454412
Ion Ratio Lower Upper
98 100
100 63.1 51.9 77.9



#51
4-Methyl-2-Pentanone
Concen: 0.344 ug/l
RT: 8.592 min Scan# 1231
Delta R.T. 0.012 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 43 Resp: 1303
Ion Ratio Lower Upper
43 100
58 39.8 35.4 53.0









#52
Toluene
Concen: 0.263 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 92 Resp: 1667
Ion Ratio Lower Upper
92 100
91 168.0 137.1 205.7



#53
t-1,3-Dichloropropene
Concen: 2.971 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 75 Resp: 1045
Ion Ratio Lower Upper
75 100
77 30.5 24.9 37.3









#54
 cis-1,3-Dichloropropene
 Concen: 2.397 ug/l
 RT: 8.378 min Scan# 1196
 Delta R.T. 0.006 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49




Instrument :
 MSVOA_X
 ClientSampled :
 IBLK


Tgt Ion: 75 Resp: 1028
 Ion Ratio Lower Upper
 75 100
 77 25.3 25.7 38.5#
 39 42.6 35.1 52.7



#58
 2-Chloroethyl Vinyl ether
 Concen: 5.198 ug/l
 RT: 8.250 min Scan# 1175
 Delta R.T. 0.000 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49

Tgt Ion: 63 Resp: 1159
 Ion Ratio Lower Upper
 63 100
 106 31.9 23.5 35.3









#59
2-Hexanone
Concen: 0.602 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 1751
Ion Ratio Lower Upper
43 100
58 40.8 30.9 92.8



#60
Dibromochloromethane
Concen: 3.190 ug/l
RT: 9.531 min Scan# 1385
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 129 Resp: 244
Ion Ratio Lower Upper
129 100
127 79.9 38.4 115.2









#61
1,2-Dibromoethane
Concen: 0.206 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:107 Resp: 563
Ion Ratio Lower Upper
107 100
109 100.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 46.225 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 95 Resp: 162468
Ion Ratio Lower Upper
95 100
174 78.8 0.0 154.6
176 75.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.061 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:117 Resp: 339719
Ion Ratio Lower Upper
117 100
82 54.5 47.4 71.2
119 32.1 26.6 39.8



#64
Tetrachloroethene
Concen: 0.497 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:164 Resp: 1348
Ion Ratio Lower Upper
164 100
166 136.7 103.4 155.2
129 156.1 70.8 106.2#
131 112.8 72.5 108.7#









#65
Chlorobenzene
Concen: 0.389 ug/l
RT: 10.085 min Scan# 1476
Delta R.T. -0.001 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:112 Resp: 2614
Ion Ratio Lower Upper
112 100
114 36.5 25.8 38.6



#67
Ethyl Benzene
Concen: 0.363 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 91 Resp: 4304
Ion Ratio Lower Upper
91 100
106 27.9 24.6 37.0









#68
m/p-Xylenes
Concen: 0.728 ug/l
RT: 10.311 min Scan# 1513
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:106 Resp: 3301
Ion Ratio Lower Upper
106 100
91 210.2 162.8 244.2



#69
o-Xylene
Concen: 0.287 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:106 Resp: 1311
Ion Ratio Lower Upper
106 100
91 171.2 108.1 324.2





#70

Styrene

Concen: 0.309 ug/l

RT: 10.665 min Scan# 1571

Delta R.T. 0.006 min

Lab File: VX023108.D

Acq: 02 Jul 2021 13:49





Instrument :

MSVOA_X

ClientSampleId :

IBLK

Tgt Ion:104	Resp:	2305
Ion Ratio	Lower	Upper
104	100	
78	51.5	40.0 60.0
103	49.3	43.4 65.0



#71

Bromoform

Concen: 4.119 ug/l




RT: 10.799 min Scan# 1593


Delta R.T. -0.006 min

Lab File: VX023108.D

Acq: 02 Jul 2021 13:49

Tgt Ion:173	Resp:	140
Ion Ratio	Lower	Upper
173	100	
175	0.0	23.9 71.8#
254	0.0	0.1 0.1#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:152 Resp: 147094
Ion Ratio Lower Upper
152 100
115 61.2 41.1 123.3
150 154.9 0.0 349.0



#73
Isopropylbenzene
Concen: 0.392 ug/l
RT: 10.969 min Scan# 1621
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:105 Resp: 4060
Ion Ratio Lower Upper
105 100
120 28.2 13.2 39.6









#74
N-amy1 acetate
Concen: 0.211 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:	43	Resp:	927
Ion	Ratio	Lower	Upper
43	100		
70	31.4	39.0	58.6#
55	27.6	21.5	32.3
61	0.0	22.1	33.1#



#76
1,2,3-Trichloropropane
Concen: 29.245 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.159 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:	75	Resp:	86567
Ion	Ratio	Lower	Upper
75	100		
77	1.2	19.1	57.5#









#77
Bromobenzene
Concen: 0.478 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:156 Resp: 1236
Ion Ratio Lower Upper
156 100
77 150.7 82.8 248.6
158 98.4 49.1 147.4



#78
n-propylbenzene
Concen: 0.529 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 91 Resp: 6335
Ion Ratio Lower Upper
91 100
120 21.5 11.5 34.5









#79
2-Chlorotoluene
Concen: 0.449 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 91 Resp: 3327
Ion Ratio Lower Upper
91 100
126 32.1 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 0.439 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 105 Resp: 3914
Ion Ratio Lower Upper
105 100
120 50.7 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 78.644 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.061 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 75 Resp: 86567
Ion Ratio Lower Upper
75 100
53 0.0 74.5 111.7#
89 0.0 37.9 56.9#



#82
4-Chlorotoluene
Concen: 0.649 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 91 Resp: 5429
Ion Ratio Lower Upper
91 100
126 29.5 14.9 44.5









#83
tert-Butylbenzene
Concen: 0.466 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:119 Resp: 3946
Ion Ratio Lower Upper
119 100
91 63.1 27.9 83.7
134 27.0 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 0.449 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:105 Resp: 3959
Ion Ratio Lower Upper
105 100
120 45.3 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 0.549 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion:105 Resp: 5985
 Ion Ratio Lower Upper
 105 100
 134 22.5 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 0.585 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX023108.D
 Acq: 02 Jul 2021 13:49

Tgt Ion:119 Resp: 5350
 Ion Ratio Lower Upper
 119 100
 134 27.2 12.9 38.7
 91 35.9 11.7 35.0#





#87

1,3-Dichlorobenzene

Concen: 0.688 ug/l

RT: 11.975 min Scan# 1786

Delta R.T. 0.000 min

Lab File: VX023108.D

Acq: 02 Jul 2021 13:49

Tgt Ion:146 Resp: 3344

Ion Ratio Lower Upper

146 100

111 48.1 20.2 60.5





148 65.9 31.9 95.7

Instrument :

MSVOA_X

ClientSampled :

IBLK



#88

1,4-Dichlorobenzene

Concen: 0.684 ug/l

RT: 11.975 min Scan# 1786

Delta R.T. -0.067 min

Lab File: VX023108.D

Acq: 02 Jul 2021 13:49




Tgt Ion:146 Resp: 3344


Ion Ratio Lower Upper

146 100

111 48.1 19.6 58.8

148 65.9 31.6 95.0









#89
n-Butylbenzene
Concen: 0.791 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 91	Resp: 6334
Ion Ratio	Lower Upper
91	100
92	51.0 27.4 82.0
134	23.2 13.2 39.5



#90
Hexachloroethane
Concen: 3.197 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 117	Resp: 335
Ion Ratio	Lower Upper
117	100
201	57.3 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 0.540 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:146 Resp: 2608
Ion Ratio Lower Upper
146 100
111 40.1 20.8 62.2
148 63.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 2.743 ug/l
RT: 12.951 min Scan# 1946
Delta R.T. 0.006 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion: 75 Resp: 147
Ion Ratio Lower Upper
75 100
155 46.3 45.3 135.9
157 82.3 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 1.377 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:	180	Resp:	4099
Ion Ratio	Lower	Upper	
180	100		
182	88.6	47.7	143.1
145	38.6	15.0	45.0



#94
Hexachlorobutadiene
Concen: 1.899 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:	225	Resp:	2394
Ion Ratio	Lower	Upper	
225	100		
223	71.1	31.1	93.2
227	63.3	31.6	95.0









#95
Naphthalene
Concen: 2.082 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Instrument :
MSVOA_X
ClientSampled :
IBLK

Tgt Ion:128 Resp: 9084
Ion Ratio Lower Upper
128 100
127 15.7 10.3 15.5#
129 11.7 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 1.162 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX023108.D
Acq: 02 Jul 2021 13:49

Tgt Ion:180 Resp: 3480
Ion Ratio Lower Upper
180 100
182 96.9 47.6 142.9
145 37.4 16.1 48.3

