

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023115.D
 Acq On : 02 Jul 2021 18:18
 Operator : JC/MD
 Sample : M2892-03DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4DL

Quant Time: Jul 05 02:56:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	243101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	388513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	350350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	164521	50.354	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.700%
35) Dibromofluoromethane	5.397	113	123696	49.304	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	465119	49.700	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.400%
62) 4-Bromofluorobenzene	11.085	95	164479	45.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.040%
Target Compounds						
19) Methyl tert-butyl Ether	3.123	73	329711	40.681	ug/l	99
29) Tetrahydrofuran	5.032	42	1696	1.354	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023115.D
Acq On : 02 Jul 2021 18:18
Operator : JC/MD
Sample : M2892-03DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4DL

Quant Time: Jul 05 02:56:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

