

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023121.D
 Acq On : 02 Jul 2021 20:39
 Operator : JC/MD
 Sample : M2924-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 05 02:58:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	231870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	381695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	343717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145613	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	164030	52.636	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.280%
35) Dibromofluoromethane	5.397	113	120733	48.982	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.960%
50) Toluene-d8	8.653	98	454309	49.412	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.820%
62) 4-Bromofluorobenzene	11.085	95	162835	45.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023121.D
Acq On : 02 Jul 2021 20:39
Operator : JC/MD
Sample : M2924-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 05 02:58:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

