

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023125.D
 Acq On : 02 Jul 2021 22:14
 Operator : JC/MD
 Sample : M2898-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-03-6.5-11.5-062921

Quant Time: Jul 05 02:59:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	231737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	383584	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	344149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144326	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	163654	52.545	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.100%
35) Dibromofluoromethane	5.397	113	122926	49.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.260%
50) Toluene-d8	8.653	98	455183	49.263	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.085	95	163257	45.767	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.540%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4847	3.934	ug/l	90
30) Chloroform	5.111	83	3049	0.660	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023125.D
Acq On : 02 Jul 2021 22:14
Operator : JC/MD
Sample : M2898-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-03-6.5-11.5-062921

Quant Time: Jul 05 02:59:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

