

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023127.D
 Acq On : 02 Jul 2021 23:01
 Operator : JC/MD
 Sample : M2907-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02-063021

Quant Time: Jul 05 02:59:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	240688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	398430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	167873	51.895	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.800%	
35) Dibromofluoromethane	5.397	113	125465	48.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.520%	
50) Toluene-d8	8.653	98	471276	49.104	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.200%	
62) 4-Bromofluorobenzene	11.085	95	171669	46.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.660%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2608	2.038	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023127.D
Acq On : 02 Jul 2021 23:01
Operator : JC/MD
Sample : M2907-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02-063021

Quant Time: Jul 05 02:59:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

