

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023128.D
 Acq On : 02 Jul 2021 23:24
 Operator : JC/MD
 Sample : M2907-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-063021

Manual Integrations
 APPROVED

SAM
 7/6/2021 1:51:34 PM

Quant Time: Jul 05 03:00:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	235659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	385963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	346826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	160293	50.609	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.220%			
35) Dibromofluoromethane	5.397	113	122058	48.972	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.940%			
50) Toluene-d8	8.653	98	462950	49.795	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.580%			
62) 4-Bromofluorobenzene	11.085	95	162574	45.295	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.580%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	18625	14.864	ug/l #	83
20) Methylene Chloride	2.794	84	4959	1.838	ug/l	97
25) 2-Butanone	4.587	43	5342	2.827	ug/l	93
29) Tetrahydrofuran	5.038	42	2046m	1.684	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023128.D
Acq On : 02 Jul 2021 23:24
Operator : JC/MD
Sample : M2907-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-063021

Manual Integrations
APPROVED

SAM
7/6/2021 1:51:34 PM

Quant Time: Jul 05 03:00:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

