



Bromomethane



Response = 2.180e-001 * Amt + 2.529e-002
 Coef of Det (r^2) = 0.997634 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

t-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 5.891\text{e-}001 * \text{Amt} - 3.224\text{e-}002$$


Coef of Det (r^2) = 0.998724 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

cis-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 6.061\text{e-}001 * \text{Amt} - 2.634\text{e-}002$$


Coef of Det (r^2) = 0.999230 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

2-Chloroethyl Vinyl ether


Response Ratio



Response = $3.014 \times 10^{-1} \times \text{Amt} - 2.827 \times 10^{-2}$
Coef of Det (r^2) = 0.999635 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

Dibromochloromethane

Response Ratio



$$\text{Response} = 4.188\text{e-}001 * \text{Amt} - 2.607\text{e-}002$$


Coef of Det (r^2) = 0.998375 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

Bromoform

Response Ratio



$$\text{Response} = 3.306\text{e-}001 * \text{Amt} - 2.682\text{e-}002$$


Coef of Det (r^2) = 0.997077 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

trans-1,4-Dichloro-2-butene

Response Ratio



$$\text{Response} = 3.983\text{e-}001 * \text{Amt} - 3.800\text{e-}002$$


Coef of Det (r^2) = 0.998800 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

Hexachloroethane

Response Ratio



$$\text{Response} = 5.654\text{e-}001 * \text{Amt} - 3.387\text{e-}002$$


Coef of Det (r^2) = 0.998294 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

1,2-Dibromo-3-Chloropropane

Response Ratio



$$\text{Response} = 2.894\text{e-}001 * \text{Amt} - 1.488\text{e-}002$$


Coef of Det (r^2) = 0.998951 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021

Naphthalene

Response Ratio



$$\text{Response} = 3.830\text{e}+000 * \text{Amt} - 9.772\text{e}-002$$

Coef of Det (r^2) = 0.999425 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M

Calibration Table Last Updated: Fri Jul 02 14:29:42 2021