

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
 Data File : VX023118.D
 Acq On : 02 Jul 2021 19:29
 Operator : JC/MD
 Sample : M2924-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 05 02:57:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	231534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	383220	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127854	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	161755	51.981	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.960%
35) Dibromofluoromethane	5.397	113	121822	49.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.460%
50) Toluene-d8	8.653	98	456447	49.447	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.085	95	154251	43.284	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070221\
Data File : VX023118.D
Acq On : 02 Jul 2021 19:29
Operator : JC/MD
Sample : M2924-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 05 02:57:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 231534
Ion Ratio Lower Upper
168 100
99 61.0 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 51.981 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29

Tgt Ion: 65 Resp: 161755
Ion Ratio Lower Upper
65 100
67 50.8 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:114 Resp: 383220
Ion Ratio Lower Upper
114 100
63 21.7 0.0 40.8
88 17.0 0.0 32.8



#35
Dibromofluoromethane
Concen: 49.227 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29

Tgt Ion:113 Resp: 121822
Ion Ratio Lower Upper
113 100
111 103.7 81.9 122.9
192 19.2 16.7 25.1









#50
Toluene-d8
Concen: 49.447 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 456447
Ion Ratio Lower Upper
98 100
100 63.3 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 43.284 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29

Tgt Ion: 95 Resp: 154251
Ion Ratio Lower Upper
95 100
174 77.4 0.0 154.6
176 73.7 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tgt Ion:117 Resp: 332621
Ion Ratio Lower Upper
117 100
82 60.0 47.4 71.2
119 31.5 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023118.D
Acq: 02 Jul 2021 19:29

Tgt Ion:152 Resp: 127854
Ion Ratio Lower Upper
152 100
115 61.9 41.1 123.3
150 157.8 0.0 349.0

