

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042130.D
 Acq On : 02 Jul 2024 11:25
 Operator : JC/MD
 Sample : VX0702WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0702WBL01

Quant Time: Jul 03 03:30:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156389	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	240983	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75552	41.895	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.800%
35) Dibromofluoromethane	5.385	113	74637	43.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.300%
50) Toluene-d8	8.647	98	279255	46.488	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.980%
62) 4-Bromofluorobenzene	11.079	95	104374	47.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042130.D
Acq On : 02 Jul 2024 11:25
Operator : JC/MD
Sample : VX0702WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0702WBL01

Quant Time: Jul 03 03:30:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

