

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042134.D
 Acq On : 02 Jul 2024 13:01
 Operator : JC/MD
 Sample : P3143-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-TB-20240628

Quant Time: Jul 03 03:32:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136046	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	206485	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67146	42.801	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.600%
35) Dibromofluoromethane	5.385	113	65628	44.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.580%
50) Toluene-d8	8.647	98	239661	46.563	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.120%
62) 4-Bromofluorobenzene	11.079	95	87825	46.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042134.D
Acq On : 02 Jul 2024 13:01
Operator : JC/MD
Sample : P3143-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-TB-20240628

Quant Time: Jul 03 03:32:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

