

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042136.D
 Acq On : 02 Jul 2024 13:48
 Operator : JC/MD
 Sample : P3143-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-698-700

Quant Time: Jul 03 03:33:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 07/03/2024
 Supervised By :Semsettin Yesilyurt 07/03/2024

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|------------------------------|--------|----------------|----------|--------|---------|----------|
| ----- | | | | | | |
| Internal Standards | | | | | | |
| 1) Pentafluorobenzene | 5.550 | 168 | 163553 | 50.000 | ug/l | # 0.00 |
| 34) 1,4-Difluorobenzene | 6.763 | 114 | 254130 | 50.000 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 10.055 | 117 | 248169 | 50.000 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152 | 127331 | 50.000 | ug/l | 0.00 |
| System Monitoring Compounds | | | | | | |
| 33) 1,2-Dichloroethane-d4 | 5.958 | 65 | 81881 | 43.416 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range 78 - 117 | Recovery | = | 86.840% | |
| 35) Dibromofluoromethane | 5.385 | 113 | 79575 | 44.130 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range 75 - 124 | Recovery | = | 88.260% | |
| 50) Toluene-d8 | 8.647 | 98 | 293718 | 46.367 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range 92 - 112 | Recovery | = | 92.740% | |
| 62) 4-Bromofluorobenzene | 11.079 | 95 | 112379 | 48.126 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range 83 - 123 | Recovery | = | 96.260% | |
| Target Compounds | | | | | | |
| | | | | | | Qvalue |
| 9) 1,1,2-Trichlorotrifluo... | 2.318 | 101 | 7290 | 5.155 | ug/l | 88 |
| 16) Acetone | 2.392 | 43 | 4450 | 5.374 | ug/l # | 84 |
| 25) 2-Butanone | 4.605 | 43 | 1873m | 1.454 | ug/l | |
| 27) cis-1,2-Dichloroethene | 4.495 | 96 | 1302 | 0.726 | ug/l | 80 |
| 30) Chloroform | 5.104 | 83 | 939 | 0.340 | ug/l | 86 |
| 44) Trichloroethene | 7.122 | 130 | 101584 | 58.683 | ug/l | 85 |
| 64) Tetrachloroethene | 9.268 | 164 | 824 | 0.456 | ug/l # | 71 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042136.D
 Acq On : 02 Jul 2024 13:48
 Operator : JC/MD
 Sample : P3143-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-698-700

Quant Time: Jul 03 03:33:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 07/03/2024
 Supervised By :Semsettin Yesilyurt 07/03/2024

