

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042141.D
 Acq On : 02 Jul 2024 15:45
 Operator : JC/MD
 Sample : P3123-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jul 03 03:35:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148493	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232752	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75978	44.372	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.740%
35) Dibromofluoromethane	5.385	113	74432	45.069	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.140%
50) Toluene-d8	8.647	98	269169	46.394	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.780%
62) 4-Bromofluorobenzene	11.079	95	101710	47.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042141.D
Acq On : 02 Jul 2024 15:45
Operator : JC/MD
Sample : P3123-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jul 03 03:35:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

