

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042142.D
 Acq On : 02 Jul 2024 16:08
 Operator : JC/MD
 Sample : P3119-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20240627

Quant Time: Jul 03 03:35:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133581	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	208232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68547	44.501	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.000%
35) Dibromofluoromethane	5.385	113	65054	44.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.060%
50) Toluene-d8	8.647	98	241657	46.557	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.120%
62) 4-Bromofluorobenzene	11.079	95	91607	47.877	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042142.D
Acq On : 02 Jul 2024 16:08
Operator : JC/MD
Sample : P3119-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20240627

Quant Time: Jul 03 03:35:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

