

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042147.D
 Acq On : 02 Jul 2024 18:05
 Operator : JC/MD
 Sample : P3119-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17-20240627

Quant Time: Jul 03 03:37:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	134528	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	206092	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68496	44.155	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.300%
35) Dibromofluoromethane	5.385	113	65468	44.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.646	98	242622	47.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.460%
62) 4-Bromofluorobenzene	11.079	95	89382	47.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.400%
Target Compounds						
						Qvalue
16) Acetone	2.391	43	1583	2.324	ug/l	92
30) Chloroform	5.098	83	26147	11.522	ug/l	95
64) Tetrachloroethene	9.274	164	888	0.593	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042147.D
Acq On : 02 Jul 2024 18:05
Operator : JC/MD
Sample : P3119-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17-20240627

Quant Time: Jul 03 03:37:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

