

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070224\  
 Data File : VX042148.D  
 Acq On : 02 Jul 2024 18:29  
 Operator : JC/MD  
 Sample : P3119-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW16D-20240627

Quant Time: Jul 03 03:38:19 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 21 23:51:58 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 145523   | 50.000   | ug/l   | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 229382   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 221148   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 102237   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 76466    | 45.568   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 91.140%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 73674    | 45.266   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 90.540%  |
| 50) Toluene-d8              | 8.647  | 98    | 261914   | 45.807   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 91.620%# |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 95613    | 45.363   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 90.720%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          |        | Qvalue   |
| 16) Acetone                 | 2.386  | 43    | 1837     | 2.493    | ug/l   | 98       |
| 30) Chloroform              | 5.093  | 83    | 65012    | 26.483   | ug/l   | 98       |
| 64) Tetrachloroethene       | 9.275  | 164   | 1204     | 0.748    | ug/l # | 78       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070224\  
Data File : VX042148.D  
Acq On : 02 Jul 2024 18:29  
Operator : JC/MD  
Sample : P3119-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW16D-20240627

Quant Time: Jul 03 03:38:19 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061824W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 21 23:51:58 2024  
Response via : Initial Calibration

